



Level



Pressure



Flow



Temperature



Liquid
Analysis



Registration



Systems
Components



Services



Solutions

Operating Instructions

SPECTRON TP CA72TP-C/D

Analyzer for the spectrophotometric determination of total phosphorus using the molybdate-vanadate method



Table of contents

1	Safety instructions	5
1.1	Designated use	5
1.2	Installation, commissioning and operation	5
1.3	Operational safety	5
1.4	Return	6
1.5	Notes on safety conventions and icons	6
2	Identification	7
2.1	Device designation	7
2.2	Scope of delivery	8
2.3	Certificates and approvals	8
3	Installation	9
3.1	The analyzer at a glance	9
3.2	Incoming acceptance, transport, storage	11
3.3	Installation conditions	11
3.4	Installation instructions	13
3.5	Post-installation check	13
4	Wiring	14
4.1	Electrical connection	14
4.2	Signal terminals	15
4.3	Switching contacts	16
5	Operation	17
5.1	Display and operating unit	17
5.2	Local operation	17
5.3	Operation in the measuring mode	18
6	Commissioning the analyzer	20
6.1	CLR start	20
6.2	Adapting the configuration data	21
6.3	Calibration	26
7	Programming	27
7.1	SETTING programming menu	27
7.2	LISTS programming menu	28
7.3	TEST programming menu	30
7.4	PRINTER programming menu	31
8	Method	32
8.1	General information	32
8.2	Measurement methods	32
8.3	Spectrophotometric determination of total phosphorus using the molybdate-vanadate method	33
8.4	Measurement optimization	34
9	Service	36
9.1	Maintenance schedule	36
9.2	Visual inspection	37
9.3	Releasing block of pump P1 (overload protection)	37
9.4	Changing the tube of pump P1	38
9.5	Calibrating pumps P1, P3 and P4	40

9.6	Calibrating the measuring system	42
9.7	Optics chamber maintenance	45
9.8	Topping up reagent	48
9.9	Removal from service	50
10	Troubleshooting	51
10.1	Troubleshooting instructions	51
10.2	Error messages	51
10.3	Return	55
10.4	Disposal	55
11	Technical data	56
11.1	Input	56
11.2	Output	56
11.3	Power supply	56
11.4	Performance characteristics	56
11.5	Environment	57
11.6	Process	57
11.7	Mechanical construction	57
11.8	Human interface	57

1 Safety instructions

1.1 Designated use

The analyzer is a compact photometric analytical system.

It is designed for monitoring the content of phosphorus in sewage treatment plants and surface waters.

The CA72TP is particularly suited to the following applications:

- Monitoring the sewage treatment plant outlet
- Monitoring process waters
- Monitoring surface waters
- Monitoring cooling tower water

Any use other than that described here compromises the safety of persons and the entire measuring system and is therefore not permitted.

The manufacturer is not liable for damage resulting from improper or non-designated use.

1.2 Installation, commissioning and operation

Note the following points:

- Installation, commissioning, operation and maintenance of the measuring system must only be carried out by trained technical personnel.
The technical personnel must be authorized for the specified activities by the owner/operator.
- The electrical connection may only be performed by an electrical technician.
- The technical personnel must have read and understood these Operating Instructions and must follow the instructions they contain.
- Before commissioning the entire measuring point, make sure all the connections are correct.
Ensure that electrical cables and tube connections are not damaged.
- Do not operate damaged products. Secure them against unintentional commissioning. Mark the damaged product as defective.
- Measuring point faults may only be rectified by authorized and specially trained personnel.
- If faults cannot be rectified, the products must be taken out of service and secured against unintentional commissioning.
- Repairs not described in these Operating Instructions may only be carried out directly at the manufacturer's or by the service organization.

1.3 Operational safety

The analyzer is designed to meet state-of-the-art safety requirements, has been tested and left the factory in a condition in which it is safe to operate.

Relevant regulations and European standards have been observed.

As the user, you are responsible for complying with the following safety conditions:

- Installation instructions
- Local prevailing standards and regulations.

1.4 Return

If the analyzer has to be repaired, please send the analyzer *cleaned* to your sales center. Use the original packaging when returning the device.

Please enclose a duly completed "Declaration of Contamination and Cleaning" form with the packaging and shipping documents (copy the second-last page of these Operating Instructions). The device cannot be repaired if a duly completed form is not enclosed!

1.5 Notes on safety conventions and icons



Warning!

This symbol alerts you to hazards. Failure to heed the warning can result in serious damage to equipment or personal injury.



Caution!

This symbol alerts you to possible faults which could arise from incorrect operation. Failure to heed the caution can result in damage to equipment.



Note!

This symbol indicates important items of information.

2 Identification

2.1 Device designation

2.1.1 Nameplate

Compare the order code on the nameplate (on the analyzer) with the product structure and your order.




Fig. 1: Nameplate

You can find the following information on the nameplate:

- Order code (device version)
- Serial number
- Measuring range
- Outputs and communication
- Power supply
- Degree of protection
- (Permitted) ambient temperature

2.1.2 Product structure

Select one feature from each section in the following structure:

Measuring range					
	A	0.05 - 2 mg total P/1 (blue)			
	B	0.1 - 5 mg total P/1 (blue)			
	C	0.3 - 8 mg total P/1 (yellow)			
	D	0.5 - 25 mg total P/1 (yellow)			
Power supply					
	0	230 VAC 50/60 Hz			
	1	115 VAC 50/60 Hz			
Sample conditioning					
	A	Not selected			
	B	1 x PA-2; PVC; 1 - 8 m³/h wastewater			
	C	1 x PA-3; PVC; 0,1 - 1 m³/h wastewater			
Device language, documentation					
	1	German			
	2	English			
Data storage medium					
	A	Not selected			
	B	Disk drive			
	C	SD card slot			
CA72TP-					Order code

You can select more than one element from the following list. These are optional items and do not have to be ordered:

Analyzer mounting (optional, select one option only)	
E1	Wall mounting
E2	Base
Communication (optional, select one option only)	
F1	RS 232 unidirectional
F2	PROFIBUS DP
Accessories, enclosed (optional, multiple selection possible)	
H1	Maintenance kit for measuring range A, B
H2	Maintenance kit for measuring range C, D
H3	Maintenance kit for sample conditioning PA-2
H4	Maintenance kit for sample conditioning PA-3

**Note!**

To complete your order code, simply add the optional features to the end of the order code. If you have any questions, please contact your local sales office.

2.2 Scope of delivery

The scope of delivery comprises:

- 1 analyzer with power plug
- 1 accessories pack
- 1 manufacturer's certificate
- 1 set of Operating Instructions in English
- 1 set of Operating Instructions for the heater control system

2.3 Certificates and approvals

2.3.1 C€ mark

Declaration of Conformity

The product meets the legal requirements of the harmonized European standards. The manufacturer confirms the product complies with the standards by affixing to it the C€ mark.

3 Installation

3.1 The analyzer at a glance

3.1.1 Front view




Fig. 2: Analyzer, front view

- | | | | |
|---|---|----|---|
| 1 | Main switch | 9 | Sample outlet |
| 2 | Sample conditioning PA-2 (option) with screen cartridge | 10 | Container with standard solutions |
| 3 | Measuring cell ventilation | 11 | Display and operating unit |
| 4 | Solenoid valve MV1 (inlet of optics chamber) | 12 | Operating unit |
| 5 | Peristaltic pump P1 with tube bed and throttle | 13 | Optics chamber |
| 6 | Solenoid valve MV2 (outlet of optics chamber) | 14 | Reciprocating piston pump P3 |
| 7 | Solenoid valve MV3 | 15 | Reciprocating piston pump P4 |
| 8 | Solenoid valve MV4 | 16 | Container with reagent solution 1 and 2 |

Sample is transported to the device via the sample conditioning system (2). Pump P1 (5) conveys the sample into the analytical part of the device. The solenoid valve MV1 (4) redirects the sample flow into the optics chamber (13) for sample dosing. Once sufficient sample has been dosed, MV1 switches to the sample outlet (9).

The solenoid valve MV2 (6) seals the optics chamber (13) for the duration of the measurement. The reciprocating piston pump P3 (14) doses reagent 1 (16) into the optics chamber.

Once the oxidation time elapses, the reciprocating piston pumps P3 (15) dose the reagents 2 (16) into the optics chamber. On completion of the color reaction, the absorption is measured via the optical measuring unit. The absorption is then evaluated by the computing unit and shown on the display (11).

When measurement is completed, the solenoid valve MV2 (6) opens and the sample flows into the sample outlet (9).

3.1.2 Rear view

The special key supplied opens the doors on the rear of the analyzer.



Warning!

Risk of electric shock! First switch off the main switch and then open the rear of the device.



Fig. 3: Analyzer, rear view

- 1 Electronics with terminals
- 2 Spectrometer electronics
- 3 Heater control
- 4 Optical unit light

- 5 Motor of reciprocating piston pump (P3, P4, P5)
- 6 Pump control system P1
- 7 Leak detector

3.2 Incoming acceptance, transport, storage

- Make sure the packaging is undamaged!
Notify the supplier of any damage to the packaging.
Keep the damaged packaging until the matter has been settled.
- Make sure the contents are undamaged!
Notify the supplier of any damage to the delivery contents.
Keep the damaged products until the matter has been settled.
- Check that the scope of delivery is complete and matches your order and the shipping documents.
- Use the handles provided to transport the analyzer.
- Pack the device for storage and transportation in such a way that it is reliably protected against impact and moisture. The original packaging offers the best protection. Furthermore, the permitted ambient conditions must also be observed.
- If you have any questions, please contact your supplier or your local sales center.

3.3 Installation conditions

3.3.1 Design, dimensions




Fig. 4: Dimensions including sample conditioning PA-2 (option)

The wet part is provided with a front door (swivel radius 525 mm (2.07 inch)), and the electronics part is provided with a front door with a window (290 x 130 mm) (1.14 x 0.51 inch).



Note!

Please note that the device dimensions can increase on account of optional fittings (e.g. sample conditioning systems).

3.3.2 Connecting the sample line




Fig. 5: Analyzer, left side panel

All connections are made and supply lines connected on the sides of the device, and must be prepared by the customer at the point of installation.

Sample supply

- 3/8" connection on the left side panel



Note!

Additional connections can be expected if a sample conditioning system is present. Information on these connections is provided in the documentation supplied on the fittings.

Analyzer sample outlet

- Tube connection DN 4/6 mm (compression fitting) on the left side panel
- Unpressurized outlet into open channel or pipe

Freshwater supply



Note!

A freshwater connection is only necessary for the optional sample conditioning system PA-2 or PA-3.

- G3/4 connection
- Pressure from 3.0 to 7.0 bar (45 to 105 psi)

Optics chamber sample outlet

- Unpressurized outlet

3.4 Installation instructions



Warning!

Risk of electric shock! Disconnect the device from the power supply (unplug the power supply connector). The line filter, overvoltage module and main switch are still energized even when the main switch is switched off.

Proceed as follows to install the analyzer at the designated site:

1. Set up the analyzer at the designated site.
 2. Check whether the 3-way ball cock in your analyzer is closed.
 3. Connect the sample supply (siehe Kapitel 3.3.2 "Connecting the sample line").
 4. If you have a sample conditioning system, connect the freshwater connection (siehe Kapitel 3.3.2).
 5. Check that the system is leaktight:
 - Switch on the wastewater pump
 - Check the connecting tubes for leaks
 - Switch off the wastewater pump
 6. Connect the tubes for ventilation and the sample outlet.
-  **Note!**
When setting up the unit in enclosed areas, make sure there is sufficient ventilation!
When commissioning the analyzer in enclosed areas, observe the following:
- Check whether the medium that is directed through the sample conditioning system gives off toxic gases (e.g. H₂S ...).
 - The ventilation of the sample conditioning system must be ensured externally via a tube.
 - Sufficient ventilation must be ensured if servicing the sample conditioning system.
7. If you have ordered inactive reagents, produce the reagent solutions in accordance with the mixing guidelines supplied, and produce the standard solutions in accordance with chapter 9.6.2 "Producing calibration standards".
 8. Place the canisters of the (active) standard and reagent solutions in the analyzer as specified on the labeling.
 9. Connect the canisters to the appropriate tubes:

Solution	Function
Reagent 1	Pump P3
Reagent 2	Pump P4
Standard 1	Solenoid valve MV4 (left)
Standard 2	Solenoid valve MV4 (right)

10. Close the tube bed of pump P1. Make sure that the tube coming from the tension side is straight.
11. Connect the signal outputs, limit value alarms and error alarm contacts as per chapter 4 "Wiring":
 - Connect signal cable for fault messages to connector I
 - Connect signal cable for limit value alarm to connector II
 - Guide the cables through the EMC shield box on the analyzer side

 **Note!**
Make sure the cables fed in have sufficient play for you to be able to reach the analyzer from the rear at a later stage.
12. Plug the mains plug into the socket (230 V, 50/60 Hz or optionally 115 V, 50/60 Hz) or connect the power cable in accordance with see chapter 4 "Wiring".

3.5 Post-installation check

- After mounting, check all the connections to ensure they are secure and leak-tight.
- Make sure the tubes cannot be removed unless force is applied.
- Check all tubes for damage.

4 Wiring

4.1 Electrical connection



Warning!

Risk of electric shock!

- Make sure you switch off the main switch before you open the rear of the device!
- The electrical connection may only be performed by an electrical technician.
- The electrical technician must have read and understood these Operating Instructions and must follow the instructions they contain.
- **Prior to commencing** the connection work, make sure no voltage is applied at any cable.

4.1.1 Electrical connection at a glance



Warning!

Risk of electric shock! Disconnect the device from the power supply (unplug the power supply connector). The line filter, overvoltage module and main switch are still energized even when the main switch is switched off.



Note!

You must open the rear, upper door of the analyzer to reach the terminal strip. The special key of the analyzer is required for this purpose.

4.1.2 Terminal assignment



Note!

The following diagrams illustrate an example of the terminal assignment of the mains connection for your analyzer. The terminal assignment and cable colors shown can deviate from the actual terminal assignment and cable colors!

To connect your analyzer, only use the terminal assignment on the adhesive label **inside the device!**

Connection to mains




Fig. 6: Sample wiring diagram for connection to mains

- 230 V / L / N / PE / 50 Hz / 16 A, terminal contacts at line filter
- Fixed connection with cable entry Pg11 or connection with plug with grounding contact
- 230 V / 60 Hz / 16 A version available on request
- 115 V / 60 Hz / 16 A version available on request



Warning!

Ensure that the analyzer is sufficiently grounded via the mains connection.

(The following must apply: $50\text{ V} < R I_{\text{max}}$, where I_{max} is the maximum current above which the error current protection switch is triggered, and R is the resistance between the protective ground and the device ground.)

If this cannot be ensured, the device must be grounded locally on site.

Power distribution diagram




Fig. 7: Power distribution terminal assignment

Cable	Function	Relay	Type	Function
A	Main switch power distribution	1	4 A	MV 1 sample to measuring cell
1	230 V - spectrometer electronics	2	4 A	MV 2 measuring cell seal
2	230 V - power supply, pump P3	3	4 A	MV 3 sample/standard switchover
3	230 V - power supply, pump P4	4	4 A	MV 4 standard 1/standard 2 switchover
4	Option	5	4 A	MV 5 screen rinsing (option)
5	Option	6	4 A	Option
6	MV 1 sample to measuring cell	7	4 A	Heater control
7	MV 2 measuring cell seal	8	Reed	Activation, P3, cable 2A
8	MV 3 sample/standard switchover			
9	MV 4 standard 1/standard 2 switchover			
10	Standard 2			
11	MV 5 screen rinsing			
12	Option			
13	230 V - power supply, heater control			

4.2 Signal terminals

The signal outputs are located on the right-hand side panel of your analyzer:




Fig. 8: Signal outputs of the analyzer

Signal outputs

Conne- ction	Name	Function
I	Reed relay, fault message	■ Floating contact (normally closed); max. 250 mA, max. 50 V
II	Reed relay, limit value alarm	■ Floating contact (normally closed); max. 250 mA, max. 50 V
III	Activation of pump P4	
IV	Option	
40	Current output, channel 1	■ 0 or 4 mA = start of measuring range; 20 mA = end of measuring range ■ Galvanically isolated; load max. 500 Ω (normally closed)
41	Option	

Computer interface (option)

- Computer interface RS 232 (option) on request; nine-pin D-Sub connection to COM2

4.3 Switching contacts




Fig. 9: Example of electronics frame wiring diagram

Cable	Function
20	Pump control system 24 VAC
23	Relay module 24 VDC
24	Option
25	Option
29	Leak detector DI 7

Cable	Function
31	Pump control system FO 1+2
37	Option Standby (IN) DI 5
39	Relay module data
COM1	Spectrometer electronics
COM2	RS 232 (option)

Pump control system and spectrometer control system




Fig. 10: Example of wiring diagram for pump control system and spectrometer control system

5 Operation

5.1 Display and operating unit




Fig. 11: Display and operating unit

5.2 Local operation

The analyzer has three operating modes:

- Measuring mode
- Service mode
- Programming mode

The menu keys have the following functions:


Key	Menu	Function
	Measuring mode	Press this key to get to the measuring mode. In the measuring mode, the current measured value, the measured values of the last 6 hours (in the form of a load curve) and the time are shown on the display.
	Service	Press this key to perform maintenance work (e.g. exchange of reagent) on the analyzer. In the service mode, you interrupt measuring operation by selecting a service menu.
	Programming	Press this key to go to the programming mode. Once you press the programming key, you are asked to enter the "code". The four-digit code can be found on the code card supplied with your device. In the programming mode, enter the conditions of the measuring point, and the time and name of the measuring point. Furthermore, in the programming mode you can also view interruptions to operation as a result of servicing activities or other interruptions to measuring operation.
	Help	A short help text on the program item in question appears when you press this key. If maintenance work is being performed, additional help on the instructions in the maintenance program is also displayed here. Exit the help again by pressing the ? key a second time or by pressing an arrow key.

The remaining operating keys have the following functions:

Key	Name	Function
   	Arrow keys	Use the arrow keys to move the cursor  in front of the desired item to select the item. You can use the arrow pointing to the right ( key) to enter negative values for certain input parameters in the configuration data of the program. A minus sign then appears when the  key is pressed.
	Entry	Use this key to call up a menu item and start an item in the program. Always press the Enter key to confirm your entries. If performing maintenance tasks, acknowledge every maintenance step once it has been performed by pressing the  key.
	Period	Call up the current operating status of the analyzer with the period key. Information on the current operating status and measuring signals appears on the display.
	Clear	If the CLR key is pressed in the operating mode, the company logo with the device type, the EPROM program version and the device options appears on the display.

5.3 Operation in the measuring mode

The measuring process is fully automated. Manual intervention is not possible. During measuring operation, the current measured value and the measured values of the last 6 hours are shown as line graphics on the display. The time is shown in the top right-hand corner. Shortly after commissioning, the display briefly displays "NO VALUE" before normal measuring operation commences. Fault messages are also shown on the display. They appear in the upper third of the screen.



The current absorption measurement is always shown regardless of the current analyzer operating status:
In the example above, the absorption shown corresponds to the absorption in color measurement. The level corresponds to the measured value of 0.32 mg/l total-P.

5.3.1 Scaling the line graphics

The measuring curve is scaled on the screen. The scales of the display can be adjusted in the PROGRAMMING/SETTING /SCALE menu in line with the measuring ranges to be expected.

5.3.2 Recording mode

Press the  key during measuring operation to enter the "recording mode". With the arrow keys, scroll through the measured values recorded over the last 10 days. The load curve, date and time are displayed.


-  1 day earlier
-  1 day later
-  2 hours earlier
-  2 hours later

Once you have selected a period or a corresponding load curve which you want to view in greater detail, press the  key.

5.3.3 Zoom function

The zoom function is activated in the recording mode by pressing an arrow key.

The load curve, measured value, date and time of the selected period are displayed. A small "zoom" cursor appears on the display and indicates your position in the load curve. Use the "zoom" cursor to view the selected 6-hour period.



- | | |
|---|--|
|  1 hour earlier |  2 minutes later |
|  1 hour later |  switches off the zoom function |
|  2 minutes earlier |  return to measuring operation |

6 Commissioning the analyzer

6.1 CLR start

You must perform a CLR start the first time you commission your analyzer. Here, the analyzer software is loaded with the factory settings and any changes made previously are deleted.

Proceed as follows to perform a CLR start:



Note!

There should not be any disk in the disk drive (option) when you press the CLR key!

1. Press the CLR key and hold it down for approx. 5 seconds. Switch on the main switch at the same time.

The application is loaded from the ROM.

2. Release the CLR key again.

The initial program screen appears on the display with the company logo and the program number of the software loaded.




Fig. 12: Program startup displayed

3. Press the  key.
A diagram with specifications for the tap position of the device appears on the display. Only confirm this point once you have completed all the tasks displayed in the diagram.

Following the CLR start, you have to adapt the analyzer to the conditions of your measuring point. Read chapter "Measurement optimization" for this purpose. In the programming menu, then adapt the configuration data of your analyzer to suit your measuring point.

6.2 Adapting the configuration data

1. Press the  key.
You are asked to enter the "code" for the measuring point. The four-digit code can be found on the code card supplied with your device.
2. Enter the code and then press the  key.
The following menu appears on the display:




Fig. 13: PROGRAMMING menu

During initial commissioning, you should go to all the submenus of the SETTING menu and adapt all the parameters to suit your measuring point:

6.2.1 Setting the date and time

1. Enter the SET CLOCK menu by pressing the keys:
  SETTING - SET CLOCK 
2. Using the arrow keys, move the cursor  in front of the value to be entered.
3. Enter the new date using the number pad and the period key.

DATE	:	mm.dd.yyyy
		05.15.2009
TIME	:	hh.mm.ss
		09.01.38
PLEASE TAKE NOTE OF INPUT FORMAT		

Fig. 14: Setting the time and date

4. Enter the exact time and press the  key to confirm.
The seconds indicator is displayed once the entry is confirmed.
5. Press the  key again to confirm.
The cursor returns immediately to the SETTING/RANGE DATA menu. An error message is displayed if you make a mistake when entering the data. Should this occur, enter the data again as described above.

6.2.2 **Configuring the measuring range data**

Here you can adapt the measuring-specific settings of your analyzer to suit your measuring point.

- 1. Enter the RANGE DATA menu by pressing the keys:

→ SETTING - RANGE DATA E

RANGE DATA		
CALIBRATION/DAY	:	1.00
SCREEN FLUSH/DAY	:	1.00
DAYBREAK	:	0.00
RANGE	:	2.00
OPERATION MODE 0/1/2	:	0.00
PAR 1 SCALE	:	30.00
STANDARD 1	:	1.00
STANDARD 2	:	0.00

Fig. 15: Configuring the measuring range data

- 2. Using the arrow keys, move the cursor ■ in front of the value to be changed.
- 3. Enter the new value and press the E key to confirm.
Example:
For calibration n day = 1.00, press the "1" key and then press the E key.
- 4. If you do not want to change a value, press the E key directly.

Here you can find information on all the parameters in the RANGE DATA menu:

Menu item	Parameter	Description
RANGE DATA		
	CALIBRATION/DAY	Activates automatic calibration(s) per day; Basic setting: 1.00 (1 calibration per day) Entry options: max. = 4.00 (4 calibrations per day), min. = 1.00 (1 calibration per day)
	SCREEN FLUSH/DAY	This parameter is only required when using sample conditioning systems. These are available as an option. Information on this parameter is provided in the Operating Instructions of the sample conditioning system.
	DAYBREAK	Time when the day changes. This is needed for calculating maximum values, minimum value and mean values, for creating the daily report and determining the time when automatic events start, e.g. automatic calibration. Basic setting: 0.00
	RANGE	Defines the measuring range (see chapter 8 "Method"); Basic setting: 2.00 Entry options: 1.00, 2.00 or 3.00; Range 3 only on request
	OPERATION MODE 0/1/2	Deactivates/activates parameters 1 and 2 on analyzers with two measuring systems: Basic setting: 0.00 (parameter 1 or 2) Entry options: 0.00, 1.00 or 2.00; 1 and 2 cannot be used for CA72TP-*
	PAR 1 SCALE	For selecting the scale end value of the display, printout and 0/4 mA signal output; here, enter the maximum concentration in mg/l total-P/1 that occurs at your measuring point. Basic setting: 30.00
	STANDARD 1	Concentration information of standard solution 1 in mg/l total-P; Basic setting: 1.00
	STANDARD 2	Concentration information of standard solution 2 in mg/l total-P; Basic setting: 10.00

6.2.3 Configuring the limit values

Here, you can define the limit values that you want to monitor at your measuring point. The collective alarm contact is opened when an alarm is triggered.

- Enter the ALARM LIMITS menu by pressing the keys:

→ ↓ ↓ ↓ SETTING - ALARM LIMITS E

Here you can find information on all the parameters in the ALARM LIMITS menu:

Menu item	Parameter	Description
ALARM LIMITS		
	DELAY (sec)	Amount of time in seconds the system waits before a limit value alarm is activated; Basic setting: 0.00
	PAR 1 UPPER LIMIT	Limit value for the alarm that signals the value is overshoot in mg/l total-P; Basic setting: 8000.00
	PAR 1 LOWER LIMIT	Limit value for the alarm that signals the value is undershot in mg/l total-P; Basic setting: 0.00
	SLOPE ALARM / 2 MIN	Limit value for the slope alarm in mg/l total-P; The alarm is triggered if the measured value increases by more than this value from one measuring point to the next measuring point (within 2 min). Basic setting: 8000.00

6.2.4 Configuring the basic data

In many analyzers, it is absolutely essential to adapt the basic data after a CLR start. Enter the values specified in the chapter "Method".

- Enter the BASIC DATA menu by pressing the keys:

→ ↓ ↓ ↓ SETTING - BASIC DATA E

Here you can find information on all the parameters in the BASIC DATA menu:

Menu item	Parameter	Description
BASIC DATA		
	METHOD	Definition of the measuring method (see chapter "Method"); No adjustment needs to be made for commissioning. Basic setting: 1.00
	Q P1 [ml/min]	Delivery rate of pump P1 in ml/min during measuring operation; No adjustment needs to be made for commissioning. Basic setting: 5.00
	MEASURING TIME	Time in seconds for the color reaction in measuring operation; For commissioning, accept the values recommended in chapter "Method". Basic setting: 180.00
	CALIBRATION TIME	Time in seconds for the color reaction in the calibration for the standard concentration in question; Basic setting: 180.00
	MEAS. DELAY MIN	Shortest interval in seconds between two measuring cycles; Basic setting: 120.00
	MEAS. DELAY MAX	Longest interval in seconds between two measuring cycles; Basic setting: 1200.00
	THRESHOLD MB [%]	Percentage difference between two consecutive measured values which causes the lower REACTION TIME MEAS. to be selected if the value is exceeded. Otherwise the MEAS. DELAY MAX is activated to save reagent. Basic setting: 20.00

Menu item	Parameter	Description
	T-FLUSH [sec]	This parameter is only required when using sample conditioning systems. These are available as an option. Information on this parameter is provided in the Operating Instructions of the sample conditioning system.
	EXCHANGE TIME	Time in seconds for replacing the sample in the sample tubes; Basic setting: 10.00
	OFFSET PAR 1	Correction value to offset deviations in the photometric measurement result. Negative values can also be entered. The offset is subtracted from the result determined. Basic setting: 0.00
	DC OUT 0/4 - 20 mA	Sets the signal output to 4-20 mA or 0-20 mA. Basic setting: 4.00 Entry options: 0.00 or 4.00

6.2.5 Name of the measuring site

- Enter the MEASURING SITE menu by pressing the keys:

→ ↓ ↓ ↓ ↓ SETTING - MEASURING SITE E

For identification purposes, the measuring site can be assigned a name with a maximum of 24 characters, or a maximum of 4 characters for disk recording. To enter the name of the measuring site, use the arrow keys ↑ ↓ to select the letters or characters. Use the arrow keys ← → to move the cursor to the next character. You can give a name to the files for disk recording under FILE XXXX.

6.2.6 INITMODEM menu

This menu is not relevant for controlling the CA72TP-*.

6.2.7 Documentation of all the data configured

Once you have entered all the configuration values, document your entries in the blank form on the next page.



Note!
Copy this form before you enter your values. You can then continue to document your entries if you make changes to the settings at a later stage.

Enter the values and data configured for commissioning here:

Date:			
REACTION PARAMETERS (Special menu, see chapter 8.4)		PUMP PARAMETERS (Special menu, see chapter 8.4)	
V.OPTICS CHAMBER		Q P1 100%	
SAMPLE PORTION		Q P3 / 15 STROKES	
REAG 1 PORTION		Q P4/ 15 STROKES	
REAG 2 PORTION			
PATH LENGTH			
REACTION TIME			
RANGE DATA		BASIC DATA	
CALIBRATION/DAY		METHOD	
SCREEN FLUSH/DAY		Q P1 [ml/min]	
DAYBREAK		MEASURING TIME	
RANGE		CALIBRATION TIME	
OPERAT. MODE 0/1/2		MEAS. DELAY MIN	
PAR 1 SCALE		MEAS. DELAY MAX	
STANDARD 1		THRESHOLD MB [%]	
STANDARD 2		T-FLUSH [sec]	
		EXCHANGE TIME	
		OFFSET PAR 1	

6.3 Calibration

Once all the parameters have been configured, the analyzer must be calibrated. The reagent solutions prepared in accordance with the specifications in the chapter "Producing calibration standards" are required for this purpose.

1. Go to the service menu .
2. Select the REAGENT EXCHANGE menu:
    REAGENT - EXCHANGE 
3. Using key "3" or "4", vent the reciprocating piston pump(s) P3 and P4. Comply with the protective measures (see chapter 9.8.2 "Changing the reagent").
4. Start a calibration after about one hour. To do so, go to the service menu:

   MEAS. SYSTEM - CALIBRATION 

A calibration is started.

The spectrometer aligns itself automatically to the optical conditions.

The analyzer goes to the measuring mode automatically following this calibration.



Note!

- Users are recommended to perform another calibration after measuring operation has been in progress for approx. one hour. This is because the overall system will have been run in at this stage and the heating will have caused the optics chamber to reach operating temperature.
- The analyzer is precalibrated on leaving the factory!

7 Programming

You should adapt the parameters of your analyzer to the conditions at your measuring point to:

- Maximize the effectiveness of your analyzer
- Get the right measuring results
- Minimize reagent consumption

Proceed as follows to go to the programming menu:

1. Press the  key.
You are asked to enter the "code" for the measuring point. The four-digit code can be found on the code card supplied with your device.
2. Enter the code and then press the  key.

The following menu appears on the display:




Fig. 16: *SETTING* menu

7.1 SETTING programming menu

You have to switch to the SETTING programming menu to adapt your analyzer parameters to the conditions of your measuring point:

  SETTING - RANGE DATA 



Note!

The analyzer may only be programmed by individuals specially appointed to perform such work (e.g. shift manager, production manager)!

More information on this menu is provided in chapter 6.2 "Adapting the configuration data".

7.2 **LISTS programming menu**

Enter the LISTS menu by pressing the keys:

   PROGRAMMING - LISTS 




Fig. 17: LISTS menu

In this menu, you can print configuration data, measured values, load curves and reports if a printer is connected (option) or show this information in part on the display:

Menu item	Parameter	Description
LISTS	INPUT DATA	Logs the current RANGE DATA, BASIC DATA and ALARM LIMITS
	CURVES/RECORDS:	Logs selected daily load curves and reports from the picklist (14 days)
	TODAY'S RECORD	The daily report is printed out when a new measurement day commences (from day changeover to day changeover).
	MAX MIN AVERAGE	Logs the maximum, minimum and average measured values for the 14 days saved
	MAINTENANCE RECORD	Log of all maintenance operations sorted by activity; select the operations to list them on the display. If a printer is connected, the information is printed out directly. Both the time and the date of the event are listed.
	PROGRAMME STARTED	Logs the date and time of the CLR start
	CHANGE DATA	Logs the date and time when configuration data are changed
	PUMP 1 REPLACE TUBE	Logs the date and time when the tube of pump P1 is changed
	PUMP 2 REPLACE TUBE	Not relevant for CA72TP-C/D
	CAL. OPTICS	Logs the date and time when the spectrophotometric measurement was calibrated; In the case of Method 4, the axis intercept X0 (offset) and the slope of the calibration line (see chapter 9.6 "Calibrating the measuring system") are documented, as are the absorption values for standard solution 1 and 2 (on a separate line).. Please note that failed calibrations are indicated by an asterisk (*). In such situations, the system continues to work with the old calibration values.
	CAL. PUMPS	Logs the date and time of a pump calibration. Furthermore, the calculated delivery volume and the pump number (1, 3, 4) are documented on a second line.
	OPTICS CHAMBER	Logs the date and time when the corresponding point in the service menu is selected.

Menu item	Parameter	Description
	SCREEN FLUSH	This parameter is only required when using sample conditioning systems. These are available as an option. Information on this parameter is provided in the Operating Instructions of the sample conditioning system.
	BYPASS-SCREEN	Logs the date and time when the corresponding point in the service menu is selected.
	REAGENT EXCHANGE	Logs the date and time when the corresponding point in the service menu is selected.
	STANDBY	Option for accessories
	SETTING OPTICS	Logs the duration of a spectrometer analysis in milliseconds and the intensity found during zero measurement in the calibration; This information is only important for service technicians.
	CAUSTIC FLUSH	Logs the date and time of alkaline rinsing.
	CAL. ELECTRODE CHAMBER	This parameter is only relevant if using analyzers with two measuring systems. Information on this parameter is provided in the Operating Instructions of the analyzer.
	ALARM RECORD	Logs all the alarms with the date and time of the event. Select this list to view it on the display or print it out if a printer is connected.
	POWER CUT	Logs when a power failure took place.
	POWER ON	Logs when the power was reestablished.
	OUT OF RANGE ON	Logs when the measuring range of the analyzer is exceeded. Please note that this can also include incorrect calibrations or lack of reagent.
	OUT OF RANGE OFF	Logs when overranging stopped; "NO VALUE" appears on the display for the period between OUT OF RANGE ON and OUT OF RANGE OFF".
	LECKAGE	Logs a leak in the device; the device switches automatically to a standby state.
	SHOW COMPLETE RECORDS	Displays all the saved events in chronological order; the last 200 events are saved in the list.
	RECORD ON DISK	Saves selected daily load curves and reports to a disk; the measured data for the last 10 days can be called up from the analyzer.
	DISK ERROR	Logs errors when saving data to a disk (data storage medium missing, write protection activated, not enough storage space etc.)

7.3 TEST programming menu

The TEST programming menu contains test programs for testing the function of the analyzer. Enter the TEST menu by pressing the keys:




Fig. 18: TEST menu



Note!
Before running the test programs select the MEASUREMENT OFF parameter to ensure an alarm is not triggered.

Menu item	Parameter	Description
TEST	TEST OF INPUTS	
	OPTICS	Test program for the optical measuring system. Here you can see what absorption value the spectrometer sends the analyzer for analysis. If you use this menu item without selecting MEASUREMENT OFF beforehand, measuring operation continues in the background. You are then able to view the analysis for all three methods simultaneously for the current range (see setting for RANGE DATA). This helps you decide what method is best for your measuring point.
	DIGITAL INPUTS	Displays the switching states of the switch inputs
	TEST OF OUTPUTS	
	DC-SIGNAL	Sets the analog current outputs to any value between 0 and 20 mA
	PUMPS	Parameter for testing the function of the pumps (see chapter 9)
	DIGITAL OUTPUTS	Functional test of the switch outputs; displays the switching states of the switch outputs
		<u>Assignment of the switch outputs:</u>
		1 = MV1 Sample in cell OFF Sample past cell ON
		2 = MV2 Measuring cell closed OFF Measuring cell open ON
		3 = MV3 Sample OFF STANDARD ON
		4 = MV4 Standard 1 OFF Standard 2 ON
		5 = MV5 Option
		6 = MV6 Option
		7 = Heater OFF ON
		8 = Pump P3 switch between ON/OFF = 1 pump stroke
		9 = Fault message
		10 = Limit value alarm
		11 = Pump P4 switch between ON/OFF = 1 pump stroke
		12 = Option
	TEST FREQUENCY I/O	Tests the inputs and outputs to check the function of the I/O card; to test the frequency inputs and outputs, the outputs can be set to a certain frequency and be read out from one input after bridging the connections.

Menu item	Parameter	Description
	TEST COM	Displays the transmission data for the RS232 interface (option)
	MEASUREMENT OFF	Disables measuring operation; the MEASUREMENT OFF operating status is displayed to perform tests without triggering an alarm.
	PRINTER	Optional

7.4 PRINTER programming menu

Enter the PRINTER menu by pressing the keys:

     PROGRAMMING - PRINTER 

This menu is only relevant for the optional printer (accessory). Information on the PRINTER programming menu is provided in the Operating Instructions of the printer.

8 Method

8.1 General information

Phosphoric compounds can occur in natural bodies of water and wastewater in both a dissolved and undissolved state, and can be determined if the sample is pretreated appropriately. The total phosphorus content is determined following previous digestion with an oxidizer and subsequent determination as an orthophosphate.

The CA72TP-C/D analyzer uses the "molybdate-vanadate method", which is particularly sensitive to high phosphate content.

8.2 Measurement methods

Light passing through the measuring cell and the sample can be absorbed by the sample (the beam of light becomes weaker). If the light is absorbed in the visible range of the spectrum (380 nm – 780 nm), the sample exhibits a coloration that is also visible to the human eye. Substances that color the sample can be quantified on the basis of how intense the coloration is. The more intense the coloration, the greater the absorption of light.

The diode array spectrometer found in the measuring device measures the absorption of the probe at a wavelength range between 380 nm and 780 nm. According to Lambert Beer's Law, the concentration of a substance causing coloration is proportional to the absorption caused by the sample:

$$\text{Absorption} = \text{concentration} \times \text{constant of proportionality} \\ \text{(Lambert Beer)}$$

Since most substances absorb differently at different wavelengths (different constants of proportionality), one can distinguish a substance from other absorbing substances using the special "form" of the absorption spectrum (all absorptions at all wavelengths beside one another) of a substance. As a spectrometer is used, the analyzer knows the entire absorption spectrum. This is taken into consideration when analyzing the measuring results in order to clear interference influences.

The SPECTRON TP CA72TP-C/D uses the following methods to analyze the absorption:

Method 1

This method analyzes the difference between two absorptions at different wavelengths.

Method 1 is selected as the standard method. It returns good results and is not affected by sample turbidity.

Method 2

The first derivative of the absorption spectrum is used for analysis purposes. This method can be selected as an alternative to Method 1 if poor results are obtained with the other methods on account of difficult wastewater situations (turbidity, inherent color, etc.). This method is not recommended for every wavelength, however (see the special notes). A slightly larger variation in measured value should be expected.

Method 3

This method uses the second derivative of the absorption spectrum for analysis purposes. Under certain circumstances, this method can return better results with colored water than the other two methods.

Please note, however, that this method is also not recommended for every wavelength (see special notes) and that the measured value noise is slightly higher than for Method 1.

8.3 Spectrophotometric determination of total phosphorus using the molybdate-vanadate method

Phosphoric compounds are digested in a boiling acidic solution with the help of sodium peroxodisulfate. The peroxodisulfate ions act as the oxidizing agent here. The acid supports the hydrolysis of the polyphosphates. The digestion of the sample produces orthophosphate ions. These create a yellow phosphomolybdate complex in the highly acidic medium. The vanadium in the reagent reacts with this complex to form an intensely yellow vanadomolybdophosphoric acid. The absorption is measured at a wavelength between 380 and 480 nm.

8.3.1 Specification of the measurement

Range	Measuring range 1 ¹⁾	Measuring range 2 ²⁾	Method 1	Method 2	Method 3
1	0.2 – 16.0 mg/l total-P	0.1 – 8.0 mg/l total-P	Recommended	Do not select	From 5 mg total-P/ ¹⁾
2	1.0 – 8.0 mg/l total-P	0.5 – 40 mg/l total-P	Recommended	Recommended	Recommended
3	2.0 – 100 mg/l total-P	1.0 – 50 mg/l total-P	Only on request		

1) With path length = 10 mm

2) With path length = 20 mm

Ratio of reagent to wastewater:	0.8 parts reagent 1 + 2.2 parts reagent 2 + 7 parts wastewater With measured values under 8 mg/l total-P ¹⁾ : half the amount of reagent + 8.5 parts wastewater With measured values over 65 mg/l total-P ¹⁾ : 0.5 parts reagent 1 + 3.2 parts reagent 2 + 6 parts wastewater
Detection limit:	0.1 mg/l total-P (for path length = 20 mm and range 1)
Method variation coefficient:	5% (Method 1)
Reagent consumption:	9.75 ml per day (with 6 measurements/hour) with measured values below 8 mg/l total-P
T ₉₀ time:	Depends on the duration of the measuring cycle; 90% of the full scale value is reached after the 2nd measurement ²⁾
Shortest measuring cycle:	25 min

1) Basic device setting needs to be configured, see chapter 8.4 "Measurement optimization"

2) Plus a dead time caused by the sample conditioning of the individual analyzer

8.3.2 Interference

Primary amines and high concentrations of silicate are also determined and result in higher readings. High concentrations of organic matter or chloride can consume the oxidizing agent and result in lower readings.

The following do not affect the reading, provided the concentration does not exceed the value indicated in the table:

Concentration [mg/l (ppm)]	Ions or interference
10 000	SO_4^{2-}
1 000	Cl^-
500	Na^+ , K^+ , Ca^{2+}
50	CO_3^{2-} , NO_3^- , Fe^{2+} , Fe^{3+} , Zn^{2+} , Cu^{2+} , Ni^{2+} , Cr^{3+} , Co^{2+} , Hg^{2+}
25	Sn^{2+}
10	Pb^{2+}
5	Ag^+
0.5	Cr^{6+} , can be eliminated by increasing the amount of ascorbic acid added

8.4 Measurement optimization

8.4.1 General information

It is essential to calibrate the analyzer to obtain precise measurement results. Since the measuring task for the device turns out to be different when the device has been in operation for several weeks, the standard concentrations selected have to be reexamined. You can improve the accuracy of the measurement by choosing the right calibration standard.

Only measure as frequently as is really required. This helps you save reagent. With the MEAS. DELAY MIN, MEAS. DELAY MAX and THRESHOLD MB parameters, the analyzer also allows you reduce the frequency of measurement when there is little change in the measured value.

Example: MEAS. DELAY MIN = 120 sec
 MEAS. DELAY MAX = 600 sec
 THRESHOLD MB = 20 %

If a new measured value deviates 20% or more from the previous measured value, the system selects the short MEAS. DELAY MIN (120 sec.) as the interval between the next measurement. If the measured value deviates less than 20% from the previous value (e.g. at night), the analyzer waits 600 sec. between two measurements.



Note!

In the PROGRAMMING / TEST OF OUTPUTS / OPTICS menu, you can view the measured values for the current range for the three methods simultaneously with measurement running in the background.

You do not have to recalibrate the analyzer if you change method.

8.4.2 Special menu

In addition to configuration data in the measuring range and basic data menus, system-internal settings are entered and changed in the special menu. In contrast to all other data, the data in the special menu are retained even if a CLR start is performed.

Proceed as follows to go to the special menu:

1. Go to the programming section .
Once the code is entered, the cursor is located at the top left-hand corner.
2. Now press the "5" key three times.

REACTION PARAMETERS submenu

Pressing the "1" key takes you to the REACTION PARAMETERS submenu:

Menu item	Parameter	Description
REACTION PARAMETERS		
	V.OPTICS CHAMBER	Enter the volume of the optics chamber in milliliters here. Basic setting: 6.50 ml for 20 mm path Optional: 4.50 ml for 10 mm path
	SAMPLE PORTION REAG 1 PORTION REAG 2 PORTION	These three parameters determine the ratio for mixing the sample with reagents REG1 and REG2 . Please make sure that the sum of these numbers/parts is 100!
	PATH LENGTH	Information on the path length (10.00 mm or 20.00 mm); This cannot be changed by the customer
	REACTION TIME	Indicates (in seconds) the duration of the oxidation of the wastewater sample with reagent 1 and the standard solutions before reagent 2 is added.

Press the  key to confirm your entries. Press the  key to exit the REACTION PARAMETERS submenu.



Note!

The entries are only saved if you exit the main level of the special menu with the  key!

PUMP PARAMETERS submenu

If you are on the main level of the special menu, press the "2" key to get to the PUMP PARAMETERS submenu:

Menu item	Parameter	Description
PUMP PARAMETERS		
	Q P1 100%	Information of the delivery volume of pump P1 determined when calibrating the pump
	Q P3 / 15 STROKES Q P4 / 15 STROKES	Information on the delivery volumes of pumps P3 and P4 from pump calibration

Press the  key to confirm your entries. Press the  key to exit the REACTION PARAMETERS submenu.

MEASURING RANGE LIMIT. submenu

If you are on the main level of the special menu, press the "4" key to get to the MEASURING RANGE LIMIT. submenu:

Menu item	Parameter	Description
MEASURING RANGE LIMIT.		
	Max. Abs.B1[abs]	Maximum absorption for range 1 (standard)
	Max. Abs.B2[abs]	Maximum absorption for range 2 (optional)
	Max. Abs.B3[abs]	Maximum absorption for range 3 (optional)
	Diff.Abs[abs]	Maximum absorption difference

Press the  key to confirm your entries. Press the  key to exit the REACTION PARAMETERS submenu.

Documentation of the entries in the special menu

If an optional printer is connected, you can also print out all the entries. For this purpose, you must press the "9" key on the main level of the special menu.

9 Service

Press the  key:




Fig. 19: SERVICE menu



Caution!
Risk of injury and infection!
When performing maintenance work on the device, always wear protective gloves to prevent damage to the skin and infection through contact with wastewater.

The program guides you step by step through the maintenance tasks. Measuring operation is interrupted if you select a program item in the service menu. Perform the maintenance task and confirm each step shown on the display by pressing the  key once you have performed the step. Pressing the  key by way of confirmation takes you to the next step in the program. The measuring device returns automatically to the measuring mode once the last step displayed has been confirmed.



Note!
The PUMPS/PUMP 2 REPLACE TUBE service menu is not relevant for CA72TP-C/D. No maintenance tasks are performed on CA72TP-C/D with this menu item.

9.1 Maintenance schedule

Timeframe	Maintenance
Daily	Visual inspection (see chapter 9.2 "Visual inspection")
Every 2 weeks	Replace standard solutions (see chapter 9.8.2 "Changing the reagent")
Every 2 weeks	Replace reagents (see chapter 9.8.2 "Changing the reagent")
Every 6 weeks	Measure the capacity of pumps P1, P3, P4 (see chapter 9.5 "Calibrating pumps P1, P3 and P4")
Every 6 weeks	Check the sample drain tube and clean if necessary
Every 3 months	pump tube P1 (see chapter 9.4 "Changing the tube of pump P1")
If required	Clean the optics chamber (see chapter 9.7 "Optics chamber maintenance")

9.2 Visual inspection

Briefly check the following:

1. Are the time and date OK?
2. Is the measurement in the standard range?
3. Are the measured values plausible?



Warning!

When working with reagents, please observe the warnings on the safety data sheets. Wear protective clothing, protective gloves and protective goggles.

4. Is the sample supply line OK?
To check, place a receptacle under the 3-way cock and open the cock slightly in the direction of the receptacle.
5. Is the solenoid valve MV2 leak-tight?
Inspect the solenoid valve MV2 for any drops forming in the tube.
6. Are sufficient amounts of standard and reagent solutions available?
Make sure that sufficient amounts of reagent and standard solution are still available in the canisters.
7. Are the pump tubes OK?
Check the pump tubes for embrittlement, leaks and drop formation.
8. If a printer is available (option):
 - Is the printer connected and online?
 - Take the load curves/reports for the day out of the printer and document them in the plant operations log.

9.3 Releasing block of pump P1 (overload protection)

The peristaltic pump P1 is fitted with an overload protection system to protect it against overload. The overload protection system can be triggered if the tube bed is set too tightly or if the pump head is blocked. This causes the pump to switch off.

Eliminate the source of the problem by checking the tube bed and loosening it if necessary. Also check the pump head and loosen it if it is too tight.



Warning!

- Danger of crushing or pinching! Do not reach into the pump head when the pump is in operation!
- When working with reagents, please observe the warnings on the safety data sheets. Wear protective clothing, protective gloves and protective goggles.
- If the liquid comes into contact with the eyes or skin, wash the affected area with plenty of water and seek medical attention.
- In the case of acidic reagents, there is a risk of splashing and the danger of extreme heat generation! For this reason, never add water to the reagents!

1. Once you switch off the wastewater pump, select:
  PUMPS - PUMP 1 REPLACE TUBE 
2. Using the  key, start the peristaltic pump briefly.

If this measure did not achieve the desired result, go to the programming menu:

3.  PROGRAMMING - TEST - TEST OF OUTPUTS - PUMPS - PUMP NO. 1 

TEST PUMPS			
PUMP NO.	1 →	0 %	0.0 ml
PUMP NO.	2 →	0 %	0.0 µl

The overload switch for pump P1 is closed.

4. Reactivate analyzer operation by pressing the  key.

9.4 Changing the tube of pump P1



Warning!

- Danger of crushing or pinching! Do not reach into the pump head when the pump is in operation!
- When working with reagents, please observe the warnings on the safety data sheets. Wear protective clothing, protective gloves and protective goggles.
- If the liquid comes into contact with the eyes or skin, wash the affected area with plenty of water and seek medical attention.
- Never add water to the reagents! In the case of acidic reagents, there is a risk of splashing and the danger of extreme heat generation!



Note!

If the device has a drain valve on the optics chamber, drain the optics chamber as described below before replacing the tube.

1. Select:

  PUMPS - PUMP 1 REPLACE TUBE 

SERVICE	
PUMPS	— PUMP 1 REPLACE TUBE
MEAS.SYSTEM	— PUMP 2 REPLACE TUBE
CLEANING	— CALIBRATE PUMPS
REAGENT	

Fig. 20: PUMP 1 REPLACE TUBE service menu

2. Remove the Plexiglas disk by loosening the two knurled nuts:




Fig. 21: Loosening the knurled nuts

3. Close the 3-way cock in the direction of the sample supply line.

4. To drain the pump tube, place a receptacle under the 3-way cock and open the cock in the direction of the receptacle.

Opening the tube bed

5. Then open the tube bed throttle:



Fig. 22: Opening the tube bed throttle

6. Then open the tube bed:




Fig. 23: Opening the tube bed



Warning!

Use a twist-and-pull movement to make it easier to release the tube. This also prevents abrasions to the skin.

Releasing and removing the pump tube from the tube connection nipples

7. Turn the tube clockwise and counterclockwise to release it from the tube connection nipples:




Fig. 24: Releasing the tube connection nipples

8. Remove the tube.

Fitting the pump tube on the tube connection nipples

9. Insert the new tube into the pump.
10. Position the tube on the tube connection nipples and fit it in place by turning the tube in a clockwise direction. Once fitted, ensure that the tube **does not exhibit any twists or bends**.

Closing the tube bed

11. Push the tube bed onto the retaining bolt in such a way that the tube coming from the tension side is straight.
12. Lubricate the tube with silicone grease, if necessary, and close the tube bed throttle.
13. Check the contact pressure of the tube bed and, if necessary, correct the setting of the setscrew on the tube bed throttle.

Checking pump startup

14. Select:
  PUMPS - PUMP 1 REPLACE TUBE 
15. Using the  key, start the peristaltic pump briefly and then stop the pump again.
16. The pump head should turn smoothly, without jerks or jolts.

If the pump does not start, the setscrew of the tube bed throttle is set too tightly. Slacken the setscrew slightly.

Opening the sample supply

17. Operate the 3-way cock as indicated in the system diagram on the display.

The measuring device returns to the measuring mode and initially displays the message "OHNE MESSWERT".

9.5 Calibrating pumps P1, P3 and P4

The SPECTRON TP has a program item that performs calibration for pumps P1, P3 and P4. You require a 10 ml and 25 ml graduated beaker for this purpose.

1. Select:
     PUMPS - CALIBRATE PUMPS 




Fig. 25: PUMPS/CALIBRATE PUMPS service menu

2. First of all, the optics chamber is filled with standard solution twice to rinse out any reagent residue.

3. You are then prompted to unscrew the dosing element of the optics chamber:




Fig. 26: Optics chamber, side view

- 1 Dosing element
- 2 GL cap
- 3 Glass tube
- 4 Measuring block, optics chamber
- 5 O-ring
- 6 Graduation glasses
- 7 Light
- 8 Measuring cell outlet




Fig. 27: Front view of the measuring cell

- 9 Measuring chamber
- 10 Detector
- 11 Connection for optical waveguide
- 12 Solenoid valve
- A Valve body
- B Thread adapter nut
- C Coil housing

- Release the GL cap on the optics chamber.
- Carefully remove the dosing element of the optics chamber and hold it over the graduated cylinder.
- Press the key.

4. The following appears on the display:




Fig. 28: PUMPS/CALIBRATE PUMPS service menu

5. You can now select the pump you want to calibrate. Using the "1", "3" or "4" key, select the desired pump and follow the instructions.



Note!

You require a 25 ml graduated beaker.

If bubbles are visible in the dosing tubes, remove the bubbles by pressing the key. As soon as pump calibration is completed, the display goes to the main level.

6. Put the GL cap back on the optics chamber.
7. You can now calibrate additional pumps.

**Note!**

Always calibrate pump P1 after replacing the pump tube. Pumps P3 and P4 do not necessarily have to be calibrated.

All the pumps are calibrated at the factory so you do not have to calibrate the pumps when you put the measuring device into operation for the first time.

9.6 Calibrating the measuring system

The analyzer measuring system can be calibrated automatically or manually. The analyzer is calibrated using two standard solutions with different known concentrations of the substance to be measured.

During the calibration procedure, the two standard solutions are supplied to the system one after the other.

The measuring system measures the absorption and assigns it to the particular standard concentration. The system then uses the mathematical ratio between the concentration and measuring signal to calculate the slope and the intercept of the calibration line. Both the intercept and the slope (in relation to 1 mg total-P/l) are saved under KAL OPTIK in the maintenance list report.




Fig. 29: Analyzer calibration line

9.6.1 Determining the standard concentrations

The right choice of standard concentration is critical to the accuracy of the measurement method. Before specifying the concentrations of the standard solutions, determine the concentration range in which the analyzer should measure. The standard solutions should cover the most frequent concentrations. Please note, however, that the concentration ratio between the two standard solutions should be between 1:5 and 1:20.

If a limit value is to be monitored, it makes sense for the concentration of the limit value to correspond to the concentration of a standard solution since this guarantees maximum precision when monitoring.

**Note!**

Please note that no measurement method can measure accurately over a range greater than 1:20.

Example:

Concentration to be measured: 1 – 20 mg/l total-P
 Most frequent concentration: 10 to 20 mg/l total-P
 Limit value to be monitored: 20 mg/l total-P

2 – 20 mg/l total-P should be selected as the standard solutions here. The system can then measure accurately in the range from 2 – 20 mg/l total-P (taking the measuring range of the system into account). Higher measured value deviation can be expected below a concentration of 2 mg/l total-P and above a concentration of 20 mg/l total-P.

9.6.2 Producing calibration standards

Endress+Hauser Conducta makes different parent and standard solutions available for a range of parameters to be measured (see chapter 9.8.1). A parent solution is a concentrated solution with an exact concentration of the substance to be measured. Calibration standards can be made by suitably diluting the parent solution. You can also order ready-to-use standard solutions directly from Endress+Hauser.

To mix a calibration standard, transfer a precise quantity of parent solution into a volumetric flask of sufficient size and then fill the flask with distilled water.

The quantity of parent solution to be added can be easily calculated by applying the "Rule of Three".

Example: To create 1 liter of standard solution with a concentration of 2 mg/l total-P, take 2 ml of a parent solution with a concentration of 1000 mg/l total-P and add distilled water until the 1 liter-mark.



Note!

- Please note that standard solutions should not be used for longer than 2 weeks. Exercise caution when mixing the standard solutions.
- Any mistakes made when mixing the standard solutions have a direct impact on the subsequent measurement operation. You must always be sure that the standard prepared has the required concentration. For this reason, always work with clean vessels. In case of doubt, prepare a new standard.
- Please note that using a reference method (e.g. cell test) to check the concentration of the standard is considerably more susceptible to errors than the actual production of the standard.

For this reason, always enter the concentration which you mixed in the device. The values of the reference system are not entered. They are only used to check the plausibility of the values. If the values deviate too much from the target concentration, prepare a new standard or perform the comparative test again.

9.6.3 Entering the concentration of the standard solutions

To enter the concentration, go to the programming level and enter the concentrations under STANDARD 1 and STANDARD 2 in the RANGE DATA programming menu. Press the key to confirm your entries.

1. Select:

SETTING - RANGE DATA

RANGE DATA		
CALIBRATION/DAY	:	1.00
SCREEN FLUSH/DAY	:	1.00
DAYBREAK	:	0.00
RANGE	:	2.00
OPERATION MODE 0/1/2	:	0.00
PAR 1 SCALE	:	30.00
STANDARD 1	:	1.00
STANDARD 2	:	0.00

Fig. 30: RANGE DATA programming menu

2. Under "STANDARD 1", enter the concentration of the standard solution with the lowest concentration.
3. Press the key to confirm your entries.
4. Under "STANDARD 2", enter the concentration of the standard solution with the highest concentration.
5. Press the key to confirm your entries.

9.6.4 Starting the calibration

There are two types of calibration: manually-activated calibration and automatically-activated calibration. Use manually-activated calibration to make the system operational again after it has been cleaned or serviced.



Note!
Please note modifications to the measuring system (optics) may cause deviations in the subsequent measurements. For this reason, always activate a manual calibration after modifying the measuring system.

Manual calibration

To start the manual calibration, select:

MEAS.SYSTEM - CALIBRATION

SERVICE

PUMPS

MEAS.SYSTEM ——— **CALIBRATION**

CLEANING

REAGENT

Fig. 31: MEAS.SYSTEM/CALIBRATION service menu

Automatic calibration

The measuring device can also be calibrated automatically. This ensures that the measuring system always returns precise measurement results. At the same time, automatic calibration also checks the operational reliability of the overall system.

- 1. Select:

SETTING - RANGE DATA

RANGE DATA

CALIBRATION/DAY	:	1.00
SCREEN FLUSH/DAY	:	1.00
DAYBREAK	:	0.00
RANGE	:	2.00
OPERATION MODE 0/1/2	:	0.00
PAR 1 SCALE	:	30.00
STANDARD 1	:	1.00
STANDARD 2	:	0.00

Fig. 32: RANGE DATA programming menu

- 2. Under "CALIBRATION/DAY", enter the number of calibrations which the measuring device should perform per day. It is generally not necessary to perform more than one calibration per day.
- 3. Press the key to confirm your entries.

9.7 Optics chamber maintenance

Structure and design

The optics chamber comprises a rectangular measuring block (4) with a fitted dosing element (1), a light (7), a detector (9) and spectrometer electronics.




Fig. 33: Optics chamber, side view




Fig. 34: Front view of the measuring cell

- 1 Dosing element
- 2 GL. cap
- 3 Glass tube
- 4 Measuring block
- 5 O-ring
- 6 Graduation glasses
- 7 Light
- 8 Optics chamber outlet

- 9 Measuring chamber
- 10 Detector
- 11 Connection for optical waveguide
- 12 Solenoid valve
- A Valve body
- B Thread adapter nut
- C Coil housing


9.7.1 Cleaning the optics chamber



Note!

The measuring cell can be cleaned manually. However, this is normally not necessary.

1. If you do, however, decide to clean the optics chamber, place a large collecting vessel under the optics chamber.
2. Select:

 CLEANING - OPTICS CHAMBER 

SERVICE

PUMPS

MEAS.SYSTEM

CLEANING — OPTICS CHAMBER

REAGENT

Fig. 35: CLEANING / OPTICS CHAMBER service menu

The optics chamber is prerinsed with standard solution.



Warning!!

- When working with reagents, please observe the warnings on the safety data sheets. Wear protective clothing, protective gloves and protective goggles.
- If the liquid comes into contact with the eyes or skin, wash the affected area with plenty of water and seek medical attention.

3. Release the GL cap (fig. 34, item **2**) from the dosing element (**1**).
4. Remove the dosing element with the GL cap and screw the GL cap onto the glass sealing cap (accessory).
5. Now clean the optics chamber from above.

Cleaning the graduation glasses

Proceed as follows if you want to clean the graduation glasses in the optics chamber:

1. Release the thread adapter nut of the optical waveguide cable (**11**) on the detector (**10**) and remove the optical waveguide cable from the detector.
2. Seal the end of the optical waveguide cable with a protection cap.
3. Unscrew the detector (**10**) from the holder.
4. Remove the detector holder.
5. At the rear of the device, release the cable connection of the light.
6. Unscrew the light (**7**) from the holder and remove the light.
7. Release the light holder from the measuring block (**4**) and remove it.
8. Remove the graduation glasses (**6**) using the suction cup (accessory).



Note!

When removing the graduation glasses, make sure you do not damage the sealing O-rings (**5**). Replace them if necessary.

9. Clean the graduation glasses.

Cleaning the measuring block

10. Operate the main switch on the device and switch the device off.



Warning!!

Risk of electric shock!

11. Allow the analyzer heater to cool.



Warning!!

Danger! Hot surfaces! The analyzer heater has to cool down before cleaning is commenced!

12. Release the thread adapter nut of the solenoid valve MV2 (**B**) completely, unfasten the coil housing (**C**) and remove it from the valve body (**A**).



Caution!

Do not turn the coil housing around its axis!

13. Remove the three slotted screws located at the top of the measuring block and remove the glass tube (**3**) of the dosing element along with the O-ring (**5**) and the retaining ring.
14. Unscrew the heating sleeve and remove the heater along with the glass tube.



Warning!!

Danger! Hot surfaces! The analyzer heater has to cool down before it can be removed!

15. Now clean all the bores of the measuring block thoroughly with demineralized water and a bottle brush.
16. Also clean all the removed parts thoroughly.

Assembly

17. After cleaning, fit the glass tube (3) of the dosing element on top of the measuring block (4) together with the retaining ring and O-ring (5).
18. Press down the glass tube until the stop and tighten the three slotted screws.
19. Release the fixing sleeve.
20. Then guide the heater together with the glass tube into the bore until the stop.
21. Now move the heater sleeve together with the large O-ring to the left and screw on the sleeve fingertight.
22. Then screw on the fixing sleeve slightly.
23. Place the coil housing of the solenoid valve (C) into the valve housing (A) in such a way that the device socket is pointing to the right.

 **Caution!**
When inserting the coil housing, make sure the coil housing does not turn around its axis!
- The O-ring can be wetted slightly with demineralized water to make insertion easier.
24. Tighten the thread adapter nut (B) of the solenoid valve by hand.
25. Insert the O-rings (5).
26. Place the graduation glasses (6) into the optics chamber with the suction cup. Make sure that the suction cup is clean and does not leave any marks on the glasses.
27. Screw on the detector holder (10). Make sure the marking is pointing upwards.
28. Screw on the light holder.
29. Then screw on the detector (10).

 **Caution!**
When installing the detector make sure that the markings on the detector holder and the detector are aligned with one another.
30. Insert the optical waveguide cable (11) and tighten the thread adapter nut.
31. Check it is well sealed.
32. Screw on the light (7).
33. Now fit the connecting cable on the light and tighten the nut fingertight.
34. Unscrew the GL cap (2) of the dosing element from the safety glass.
35. Fit the dosing element (1) on the glass holder of the optics chamber and tighten it.

 **Caution!**
When inserting the dosing element, make sure that the three Teflon tubes are pointing straight towards the optics chamber and not towards the glass tube or the optics chamber body! If necessary, realign the tubes with the optics chamber.
36. Switch on the main switch again.
37. Confirm that cleaning has been performed by pressing the  key.

Manual calibration

38. Trigger a calibration manually after two to three measuring cycles (see chapter "Calibrating the measuring system").

 **Note!**
The analyzer has to be recalibrated after disassembling a component of the optical system (light, detector, optical waveguide, etc.).

9.8 Topping up reagent



Warning!

- When working with reagents, please observe the warnings on the safety data sheets.
- Wear protective clothing, protective gloves and protective goggles.
- If the liquid comes into contact with the eyes or skin, wash the affected area with plenty of water and seek medical attention.
- Never add water to the reagents! In the case of acidic reagents, there is a risk of splashing and the danger of extreme heat generation!

9.8.1 Reagents

Only use genuine containers for the measurement. We are not liable for any damage resulting from the use of other chemicals.

The following reagent components are needed to determine the total phosphorus content:

Parent solution

- 1000 mg/l (1000 ppm) $\text{PO}_4\text{-P}$; for preparing standard solutions 1 and 2
- Order numbers:
 - 1000 ml (33.8 fl.oz.): CAY248-V10C00AAE
 - 100 ml (3.38 fl.oz.): CAY248-V01C00AAE

Standard solutions ready for use, per 1 liter (33.8 fl.oz.)

- Standard 1.0 mg/l (1 ppm) $\text{PO}_4\text{-P}$; Order No. CAY242-V10C01AAE
- Standard 1.5 mg/l (1.5 ppm) $\text{PO}_4\text{-P}$; Order No. CAY242-V10C03AAE
- Standard 2.0 mg/l (2 ppm) $\text{PO}_4\text{-P}$; Order No. CAY242-V10C02AAE
- Standard 5 mg/l (5 ppm) $\text{PO}_4\text{-P}$; Order No. CAY242-V10C05AAE
- Standard 10 mg/l (10 ppm) $\text{PO}_4\text{-P}$; Order No. CAY242-V10C10AAE
- Standard 15 mg/l (15 ppm) $\text{PO}_4\text{-P}$; Order No. CAY242-V10C15AAE
- Standard 20 mg/l (20 ppm) $\text{PO}_4\text{-P}$; Order No. CAY242-V10C20AAE
- Standard 25 mg/l (25 ppm) $\text{PO}_4\text{-P}$; Order No. CAY242-V10C25AAE
- Standard 30 mg/l (30 ppm) $\text{PO}_4\text{-P}$; Order No. CAY242-V10C30AAE
- Standard 40 mg/l (40 ppm) $\text{PO}_4\text{-P}$; Order No. CAY242-V10C40AAE
- Standard 50 mg/l (50 ppm) $\text{PO}_4\text{-P}$; Order No. CAY242-V10C50AAE

Reagent set, active

- Digestion agent sodium peroxodisulfate R1, 40 g (powder)
- Color reagent R2, 2 l (67.6 fl.oz.)
- Order No. CAY249-V20AAE


Reagent 1 (sodium peroxodisulfate R1) contains the oxidizing agent and is prepared on site in accordance with the mixing instructions supplied.

Reagent 2 (color reagent) are available as a reagent pack as active and inactive reagents. The active reagent can be used directly but can only be stored for a limited period. It also has to be stored in a dark place at 4 to 8 °C (39 to 46 °F).

The inactive reagent can be stored in a dark place as per the best-before date but must be prepared following the mixing instructions before it can be used.

9.8.2 Changing the reagent

- To change the reagent, go to the service menu:

 REAGENT - EXCHANGE 




Fig. 36: REAGENT / EXCHANGE service menu

Subsequent screen:




Fig. 37: Subsequent screen

- Remove the cover from the new reagent.
- Open the cover of the canister in the device and remove together with the tube projecting into the canister.



Note!

It is advisable to have a damp cloth at hand to clean up any drops of reagent that may accidentally drip.

- Remove the empty canister.
- Put a new canister into the device.
- Connect the canisters to the appropriate tubes:

Solution	Function
Reagent 1	Pump P3
Reagent 2	Pump P4
Standard 1	Solenoid valve MV4 (left)
Standard 2	Solenoid valve MV4 (right)

- Press key "3" and "4" to convey reagent until bubble-free reagent is conveyed.

9.9 Removal from service

You must take the analyzer out of service before dispatching it or in the event of extended breaks in operation.



Warning!

- When working with reagents, please observe the warnings on the safety data sheets.
- Wear protective clothing, protective gloves and protective goggles.
- If the liquid comes into contact with the eyes or skin, wash the affected area with plenty of water and seek medical attention.
- Never add water to the reagents! In the case of acidic reagents, there is a risk of splashing and the danger of extreme heat generation!



Caution!

Prior to taking the analyzer out of service, you must rinse all the tubes thoroughly with distilled water.

Proceed as follows:

1. Switch off the wastewater pump.
2. Feed the reagent back into the canister by opening the lower connection on pump P3.
3. Connect the empty tube back to the tube connector .
4. Take a collecting vessel containing approx. 2 l of distilled water.
5. Remove the reagent canister and put the collecting vessel with the distilled water in its place.
6. To rinse the pump tubes of pump P3 and P4, select:
     REAGENT - EXCHANGE 
7. Repeatedly press keys "3" or "4" to rinse the tubes of pump P3 or P4 with the distilled water from the collecting vessel.
8. Trigger automatic cleaning of the measuring cell:
    CLEANING - OPTICS CHAMBER 
9. Clean the optics chamber manually in accordance with Kapitel "Optics chamber maintenance"
10. Release the tube bed throttle of pump P1.
11. Hold a collecting tray under the 3-way cock and seal off the supply of sample.
12. Remove the canisters.
13. Operate the main switch on the device and switch the device off.



Note!

Keep open reagents and standard solutions refrigerated. Observe the best-before date.

10 Troubleshooting

10.1 Troubleshooting instructions

The analyzer's simple design means it is not very susceptible to errors. Nevertheless, errors at the measuring point cannot be ruled out completely.

The following section thus contains possible errors, the causes behind such errors and instructions on how to rectify the errors.

10.2 Error messages

Error message on the display	Cause	Countermeasures
"Spectrometer ???"	No communication possible with the spectrometer. Transmission cables or contacts not OK.	Check the connecting cable running from COM1 to COM3 on the spectrometer.
	Power supply to the spectrometer disconnected.	Check the power supply to the spectrometer. (When the analyzer is switched on and off at the main switch, the red LED of the spectrometer flashes briefly.)
"Out of range"	An error occurred when determining the measured value. The measurement was rejected and a new measuring cycle initiated. Real measured value overshooting Error in optical path	The measuring range limit was reached. The error is displayed when the maximum absorption level permitted is overshot. Is the optical wave cable mounted? Is the LED lit? If everything is OK and the error message continues to appear on the display, switch off the measuring device and switch it back on again and trigger a calibration.

10.2.1 Implausible measured values

	Check:		If so	If not
1.	Are the values of the last calibration OK (in the normal range, no asterisks)?		2	see chapter 10.2.2 "Implausible calibration values"
2.	Is reagent still present??		4	3
3.	Top up the reagent.	Problem solved?	16	4
4.	Is the reagent conveyed free from air bubbles?		6	5
5.	Rectify defect in leak tightness in the flow.	Problem solved?	16	6
6.	Is there sample at the 3-way ball cock? Hold a beaker under the 3-way cock and turn the cock in the direction of the beaker.		8	7

	Check:		If so	If not
7.	Check the sample supply path and the wastewater pump.	Problem solved?	16	8
8.	Check the heater: Is the detector or the measuring cell warm?		10	9
9.	Check the heater, the heating regulator, relay No. 7 and the power supply.	Problem solved?	16	10
10.	Does MV1 switch? In the PROGRAMMING/TEST/TEST OF OUTPUTS/DIGITAL OUTPUTS mode, check whether MV1 switches electrically when switch output 1 is actuated.		12	11
11.	Check whether the corresponding relay, the fuse, the electrical connection and the solenoid valve are functioning correctly.	Problem solved?	16	12
12.	Is sample dripping into the measuring cell? Lift up the dosing element approx. 2 cm and check whether sample drips into the measuring cell during a new measuring cycle.		14	13
13.	Check the sample path from the 3-way cock to the dosing element and eliminate any clogging.	Problem solved?	16	14
14.	Is MV3 leak-tight? Remove the tube for standard 1 from the storage container. Standard may not be conveyed during normal operation.		16	17
15.	Replace the seal of solenoid valve MV3.	Problem solved?	16	
16.	Trigger a calibration.	Problem solved?		17
17.	Notify the Service department of Endress+Hauser Conducta GmbH&Co.KG, Division STIP. Use the service information sheet at the end of this chapter	Tel: +49 6078 786-20 Fax: +49 06078 786-88		

10.2.2 Implausible calibration values

	Check:		If so	If not
1.	Is sufficient standard in the standard containers?		3	2
2.	Top up the standard solutions.	Problem solved?	20	3
3.	Are the standard solutions conveyed free from air bubbles?		5	4
4.	Check all the seals and tubes along the delivery path.	Problem solved?	20	5
5.	Is reagent still present?		7	6
6.	Top up the reagent solutions.	Problem solved?	20	7
7.	Is the reagent conveyed free from air bubbles?		9	8
8.	Rectify defect in leak tightness in the flow of reagent.	Problem solved?	20	9
9.	Are solenoid valves MV3 and MV4 leak-tight? Is sample possibly mixing with standard, or is standard 1 mixing with standard 2? Remove the tube for standard 1 and 2 from the canister. Close the 3-way cock in the direction of the sample supply. Attach a tube with an ID of 4 mm (ID 3.2 inch) or less to the 3-way cock and place it in a beaker with water. Go to PROGRAMMING/TEST/TEST OF OUTPUTS/PUMPS and set P1 to 100%. Change to PROGRAMMING/TEST/TEST OF OUTPUTS/DIGITAL OUTPUTS and check whether only the specific sample or standard is conveyed when switch output 3 and 4 are actuated.		11	10
10.	Replace the seal of solenoid valve MV 3 and/or MV4.	Problem solved?	20	11

	Check:		If so	If not
11.	Does solenoid valve MV1 switch? In the PROGRAMMING/TEST/TEST OF OUTPUTS/DIGITAL OUTPUTS mode, check whether MV1 switches electrically when switch output 1 is actuated.		13	12
12.	Check whether the corresponding relay, the fuse, the electrical connection and the solenoid valve are functioning correctly.	Problem solved?	20	13
13.	Is sample dripping into the optics chamber? Lift up the dosing element approx. 2 cm and check whether sample drips into the optics chamber during a new measuring cycle.		15	14
14.	Check the sample path from solenoid valve MV3 to the dosing element and eliminate any clogging.	Problem solved?	20	15 Optical measurement
15.	Check the spectrometric measurement: Are the glasses of the optics chamber clean? If X0 is unusually high (print on the left-hand side of the MAINTENANCE RECORD CAL.OPTICS). Disassemble the optical wave cable and the detector (see chapter 9.7 "Optics chamber maintenance") and empty the optics chamber with the service key.		17	16
16.	Clean the graduation glasses of the optics chamber (see chapter 9.7 "Optics chamber maintenance").	Problem solved?	20	17
17.	Is the slope unusually low or high?		15	21
18.	Are the standard solutions OK? (Test with another method or prepare new standards. In doing so, please follow the instructions under Kapitel "Producing calibration standards").		21	19
19.	Replace the standard solutions with new ones.	Problem solved?	20	2
20.	Trigger a calibration (see chapter 9.6 "Calibrating the measuring system").	Problem solved?		21
21.	Notify the Service department of Endress+Hauser Conducta GmbH&Co.KG, Division STIP. Use the service information sheet at the end of this chapter.	Tel: +49 6078 786-20 Fax: +49 6078 786-88		

Service information sheet for SPECTRON TP

If problems occur, fill in this information sheet and fax it along with a description of the problem to
Endress+Hauser Conducta GmbH&Co.KG Division STIP:
Fax: +49 6078 78688

Name:		Device number:		Tel.:	
Visual inspection and trouble-shooting		Bypass and measuring cell outlet free from pressure?			
SPECIAL MENU DATA Key "1"		Key "2"			
V.OPTICS CHAMBER	=		Q P1 100%	=	If you have a printer, you can print out the data in the programming mode by pressing key "5" 3 times and then key "Q"!
SAMPLE PORTION	=		Q P3 / 15 HÜBE	=	
REAG 1 PORTION	=		Q P4/ 15 HÜBE	=	
REAG 2 PORTION	=				
PATH LENGTH	=				
REACTION TIME	=				
RANGE DATA			BASIC DATA		
CALIBRATION/DAY	=		METHOD	=	If you have a printer, you can print out the measuring range data and the basic data under PROGRAMMING/LISTS/INPUT DATA.
SCREEN FLUSH/DAY	=		Q P1[ml/min]	=	
DAYBREAK	=		MEASURING TIME	=	
RANGE	=		CALIBRATION TIME	=	
OPERATION MODE 0/1/2	=		MEAS. DELAY MIN	=	
PAR 1 SCALE	=		MEAS. DELAY MAX	=	
STANDARD 1	=		THRESHOLD MB[%]	=	
STANDARD 2	=		T-FLUSH [sec]	=	
			EXCHANGE TIME	=	
			OFFESET PAR 1	=	
Description of the measuring point			Description of the medium		
Industrial Municipal Inlet Sludge activation Secondary clarification Other (short description) Ambient temperature:			Turbidity (low, medium, high): Coloration (description of the color) Salt content mg/l: Fibrous Oily or greasy Sludge volume: Total solids: Sample temperature:		
Print out information from the programming mode/list, or note down this information on another sheet:			The alarm messages before and after the error occurred: PROGRAMMING/LISTS/ALARM RECORD The last three calibration value records PROGRAMMING/LISTS/MAINTENANCE RECORD/CAL.OPTICS. If a printer is available, print out the load curves for the 3 days before and after the error occurred.		
If available, note down the comparative measurements.			CA72TP- :		Comparative measurement:
					Performed with?

10.3 Return

If the analyzer has to be repaired, please send the analyzer *cleaned* to your sales center. Use the original packaging when returning the device.

Please enclose a duly completed "Declaration of Contamination and Cleaning" form with the packaging and shipping documents (copy the second-last page of these Operating Instructions). The device cannot be repaired if a duly completed form is not enclosed!

10.4 Disposal

The device contains electronic components and must therefore be disposed of in accordance with regulations on the disposal of electronic waste. Please observe local regulations.

11 Technical data

11.1 Input

Measured variable	Total phosphorus (TP) [mg/l]
Measuring range	■ 0.3 to 8 mg/l (CA72TP-C) ■ 0.5 to 25 mg/l (CA72TP-D)
Wavelength	380 to 480 nm

11.2 Output

Output signal	0/4 to 20 mA, galvanically isolated
Signal on alarm	Limit value alarm, fault message and two optional contacts (0.25 A / 50 V)
Load	Max. 500 Ω
Loading capacity	230 V AC, max. 2 A

11.3 Power supply

Assignment	See fig. 7
Supply voltage	230 V AC, 50/60 Hz
Power consumption	161 VA
Current consumption	0.7 A
Fuses	1 x 16 A FF (power distribution connection) 1 x 2 A T (electronics power unit)



Note!

The SPECTRON TP CA72TP is suitable for connection to industrial mains support according to EN 61326-1, class A.

11.4 Performance characteristics

Maximum measured error	±5 % of end of measuring range
Measuring interval	Approx. 2 measurements per hour (at a digestion time of 15 min.)
Time between two measurements	■ t_{meas} = sample dosing (90 s) + oxidation time (960 s) + color reaction time (180 s) + measured value calculation (180 s) + discard sample + break in measuring (optional) + rinse time (90 s) = 25 min ■ Only the oxidation time can be adjusted (0 to 3600 s)
Sample requirement	5 ml (0.17 fl.oz.) / measurement
Reagent requirement	■ Oxidizing agent REG1: 1000 ml (34 fl.oz.) / month with 25-minute measuring interval ■ REG2: 2500 ml (85 fl.oz.) / month with 25-minute measuring interval
Calibration interval	■ Selectable, 1-4 calibrations per day up to one calibration per week ■ Standard: once per day at ambient temperatures < 30 °C (86 °F)
Maintenance interval	6 months (typical)

Maintenance effort	■ Daily: visual inspection ■ Every 2 weeks: replace or top up reagents and standards ■ Every 6 weeks: clean sample conditioning system (if available) ■ Every 12 weeks: replace pump tubes and calibrate all pumps
---------------------------	---

11.5 Environment

Ambient temperature range	0 to 40 °C (32 to 100 °F)
Humidity	10 to 90 %, non-condensating
Degree of protection	IP54

11.6 Process

Sample temperature	5 to 40 °C (40 to 104 °F)
Sample flow rate	5 to 12 ml/min (0.17 to 0.4 fl.oz./min); pay attention to delivery rate of the pump!
Consistency of sample	Low level of solids, particle size < 500 ppm
Sample outlet	Unpressurized

11.7 Mechanical construction

Design, dimensions	see chapter 3 "Installation"	
Weight	Approx. 80 kg (176 lbs)	
Material	Housing Front window Valve seals Pump tubes Pump and pump seals Reagent and sample tubes Tubes for exhaust air and ventilation Discharge tubes	Aluminum, powder-coated Glass, conductive coating EPDM, PTFE EPDM, Tygon® PTFE PTFE, PFA Norprene, PE PTFE, PE

11.8 Human interface

Display and operating elements	Display LCD graphic display, 16 lines, 40 characters per line, backlit Keyboard 21 operating keys, 13 x 13 mm with pressure point
Serial interface	RS232

Declaration of Hazardous Material and De-Contamination Erklärung zur Kontamination und Reinigung

RA No.

Please reference the Return Authorization Number (RA#), obtained from Endress+Hauser, on all paperwork and mark the RA# clearly on the outside of the box. If this procedure is not followed, it may result in the refusal of the package at our facility.
Bitte geben Sie die von E+H mitgeteilte Rücklieferungsnummer (RA#) auf allen Lieferpapieren an und vermerken Sie diese auch außen auf der Verpackung. Nichtbeachtung dieser Anweisung führt zur Ablehnung ihrer Lieferung.

Because of legal regulations and for the safety of our employees and operating equipment, we need the "Declaration of Hazardous Material and De-Contamination", with your signature, before your order can be handled. Please make absolutely sure to attach it to the outside of the packaging.

Aufgrund der gesetzlichen Vorschriften und zum Schutz unserer Mitarbeiter und Betriebseinrichtungen, benötigen wir die unterschriebene "Erklärung zur Kontamination und Reinigung", bevor Ihr Auftrag bearbeitet werden kann. Bringen Sie diese unbedingt außen an der Verpackung an.

Type of instrument / sensor

Geräte-/Sensortyp _____

Serial number

Seriennummer _____

☐ Used as SIL device in a Safety Instrumented System / Einsatz als SIL Gerät in Schutzeinrichtungen

Process data/ Prozessdaten

Temperature / Temperatur _____ [°F] _____ [°C]

Pressure / Druck _____ [psi] _____ [Pa]

Conductivity / Leitfähigkeit _____ [µS/cm]

Viscosity / Viskosität _____ [cp] _____ [mm²/s]

Medium and warnings

Warnhinweise zum Medium



	Medium /concentration Medium /Konzentration	Identification CAS No.	flammable entzündlich	toxic giftig	corrosive ätzend	harmful/ irritant gesundheitsschädlich/ reizend	other * sonstiges*	harmless unbedenklich
Process medium Medium im Prozess								
Medium for process cleaning Medium zur Prozessreinigung								
Returned part cleaned with Medium zur Endreinigung								

* explosive; oxidising; dangerous for the environment; biological risk; radioactive

* explosiv; brandfördernd; umweltgefährlich; biogegefährlich; radioaktiv

Please tick should one of the above be applicable, include safety data sheet and, if necessary, special handling instructions.

Zutreffendes ankreuzen; trifft einer der Warnhinweise zu, Sicherheitsdatenblatt und ggf. spezielle Handhabungsvorschriften beilegen.

Description of failure / Fehlerbeschreibung _____

Company data / Angaben zum Absender

Company / Firma _____	Phone number of contact person / Telefon-Nr. Ansprechpartner: _____
Address / Adresse _____	Fax / E-Mail _____
Your order No. / Ihre Auftragsnr. _____	

"We hereby certify that this declaration is filled out truthfully and completely to the best of our knowledge. We further certify that the returned parts have been carefully cleaned. To the best of our knowledge they are free of any residues in dangerous quantities."

"Wir bestätigen, die vorliegende Erklärung nach unserem besten Wissen wahrheitsgetreu und vollständig ausgefüllt zu haben. Wir bestätigen weiter, dass die zurückgesandten Teile sorgfältig gereinigt wurden und nach unserem besten Wissen frei von Rückständen in gefährlicher Menge sind."

(place, date / Ort, Datum)

Name, dept./ Abt. (please print / bitte Druckschrift)

Signature / Unterschrift

For more Endress+Hauser product please visit: www.ainstru.com

Mail: Sales@ainstru.com

www.endress.com/worldwide

Endress+Hauser 
People for Process Automation