

Technical Information

GM32

In-situ Gas Analyzer,
Measuring Probe Version



Described product

Product name: GM32

Variants: GM32 Probe (measuring probe)

Manufacturer

Endress+Hauser SICK GmbH+Co. KG

Bergener Ring 27

01458 Ottendorf-Okrilla

Deutschland

Legal information

This work is protected by copyright. Any rights derived from the copyright shall be reserved for Endress+Hauser SICK GmbH+Co. KG. Reproduction of this document or parts of this document is only permissible within the limits of the legal determination of Copyright Law. any modification, shortening or translation of this document is prohibited without the express written permission of Endress+Hauser SICK GmbH+Co. KG.

The trademarks stated in this document are the property of their respective owner.

© Endress+Hauser SICK GmbH+Co. KG. All rights reserved.

Original document

This document is an original document of Endress+Hauser SICK GmbH+Co. KG.



Warning Symbols



Hazard (general)



Hazard by voltage



Hazard by unhealthy substances

Warning Levels / Signal Words

DANGER

Risk or hazardous situation which *will* result in severe personal injury or death.

WARNING

Risk or hazardous situation which *could* result in severe personal injury or death.

CAUTION

Hazard or unsafe practice which *could* result in personal injury or property damage.

NOTICE

Hazard which could result in material damage.

Information Symbols



Important technical information for this device



Important information on electric or electronic functions



Supplementary information



Link to information at another place

1	General Information	7
1.1	Responsibility of user	8
1.1.1	Designated users	8
1.1.2	Special local conditions	8
2	Product Description (Details)	9
2.1	Measuring method	10
2.1.1	Optics layout and functional principle	10
2.2	Signal evaluation	11
2.3	Zero adjust (description)	12
2.4	Reference cycle (description)	12
2.4.1	Reference cycle sequence	12
2.5	Check cycle (description)	12
2.5.1	Control elements	12
2.5.2	Check cycle (function)	13
2.5.2.1	Wavelength scale monitoring	13
2.5.2.2	Resolution drift monitoring	13
2.5.2.3	Absorption drift monitoring	13
2.5.2.4	Determining the zero and reference point	14
2.5.2.5	Monitoring the cell NO filling	14
2.5.2.6	Summary	15
2.5.3	Events and signals output	15
2.5.3.1	Check cycle signals	15
2.5.3.2	Maintenance request signal	15
2.5.3.3	Zero and reference point level output	15
2.6	Contamination compensation	16
2.6.1	Monitoring contamination	16
2.7	Automatic adjustment with alignment	16
2.8	Measuring probe in detail	17

3	CAN Connection	21
3.1	GM32 CAN connection	22
3.2	SCU connection (overview)	23
3.3	Connecting the GM32 to the CAN bus and to the SCU	24
3.3.1	Settings in GM32	25
3.3.2	Settings in the SCU	25
3.4	Mapping Table	26
4	Operation (Specialist Menus)	29
4.1	SOPAS ET (description)	30
4.2	Menu tree (in SOPAS ET)	31
4.2.1	Changing the user level	32
4.2.2	Measured values	33
4.2.2.1	Bargraph measured values	33
4.2.3	Parameters	34
4.2.3.1	Device parameters	34
4.2.3.2	Spectrometer	35
4.2.3.3	Coefficients	35
4.2.3.4	Coefficients boxmeasuring	36
4.2.3.5	Adjustment sensor	36
4.2.3.6	Logbook	37
4.2.3.7	Analog outputs	38
4.2.3.8	Analog inputs	38
4.2.3.9	Digital inputs	39
4.2.3.10	Digital outputs	39
4.2.3.11	Hardware Map	41
4.2.4	Adjustment	42
4.2.4.1	Alignment	42
4.2.4.2	Zero adjust (Service level only)	43
4.2.4.3	Boxmeasuring (Service level only)	44
4.2.5	Diagnosis	45
4.2.5.1	Device information	45
4.2.5.2	Logbook	45
4.2.5.3	Check cycle	47
4.2.5.4	Sensor values	47
4.2.5.5	Temperature unit	50
4.2.5.6	Spectra	51
4.2.6	Maintenance	56
4.2.6.1	Operating mode switch	56
4.2.6.2	Lamp hour meter	56
4.2.6.3	Save EEPROM	57
4.2.6.4	Load EEPROM	57
4.2.7	Start-up Assistant	57

4.3	Program/data structure ("Firmware", "Device description file")	58
4.3.1	Sequence of settings and file transfers	58
4.3.2	Setting the GM32 analyzer/SCU IP address	59
4.3.3	Updating the firmware (GM32 analyzer/SCU)	60
4.3.4	Loading the device description file ("Make available to SOPAS ET")	61
4.3.5	Saving and loading parameters (device data)	62
4.3.5.1	Loading parameters changed in SOPAS ET to the GM32 analyzer/SCU	62
4.3.5.2	Saving parameters (GM32 analyzer/SCU) to an external data medium	63
4.3.5.3	Loading parameters from an external data medium to the GM32 analyzer/ SCU	63
4.3.6	Loading data files from the device	63
5	Maintenance	65
5.1	Zero adjust (zero point check)	66
5.1.1	Procedure	66
5.2	Boxmeasuring	68
5.2.1	Preparing the filter box	69
5.2.2	Determining the necessary test gas concentration	71
5.2.3	Procedure	72
5.3	Test gas feeding (for GPP measuring probe)	76
5.3.1	One-off initial measurement - determining the test gas pressure	77
5.3.2	Manual test gas feeding	78
5.3.3	Using mixed gases as test gases	78
5.4	Fitting and removing the power supply unit of the connection unit	79
6	Shutting down/Restarting	81
6.1	Qualification requirements	82
6.2	Shutting down	82
6.2.1	Equipment required	82
6.2.2	Shutdown	83
6.2.2.1	Removing from the sample gas duct	83
6.2.3	Disconnecting the purge air connection	84
6.2.4	Putting the purge air unit out of operation	85
6.3	Further components on the measuring point	86
6.3.1	Connection unit and cabling	86
6.4	Conservation measures, correct storage and transport	87
6.5	Preparing for restarting	89
6.5.1	New measuring task	89
6.5.2	Previous measuring task	89
6.5.3	Purge air unit maintenance	89
7	Annex	91
7.1	Commissioning checklist	92

GM32

1 General Information

Responsibility of user

1.1 **Responsibility of user**

1.1.1 **Designated users**

This Technical Information is aimed at qualified persons familiar with the GM32 and who, based on their device-specific training and knowledge of the device as well as knowledge of the relevant regulations, can assess the tasks given and recognize the dangers involved.

1.1.2 **Special local conditions**

Follow all local laws, regulations and company-internal operating directives applicable at the installation location.

GM32

2 Product Description (Details)

Functional principle

Check cycles

Remote operation

2.1 Measuring method


2.1.1 Optics layout and functional principle

The GM32 multi-component analyzer is based on in-situ technology with opto-electronic direct measurement. Measured values are recorded without contact directly in the gas flow using a measuring probe protruding into the duct. Depending on the version, this has either an open measuring gap or a gas permeable membrane insert (filter) through which the gas mixture being measured flows - the active measuring path.

The GM32 SR-unit determines the concentration of the respective gases based on wavelength-specific light absorption by the gas mixture in the active measuring path.

Light from the sender/receiver unit (SR-unit) illuminates the active measuring path in the gas duct which is then reflected back again by a triple reflector at the end of the probe. The beam splitter diverts the returning light to the polychromatic subassembly comprising a condenser lens with slit diaphragm, optical grating and receiver element. The optical grating dissects the returning light spectrally and maps it to the receiver element, a highly sensitive diode array.

Fig. 1 Measuring principle




2.2

Signal evaluation

The optimized algorithms of the GM32 evaluation electronics process the measurement signal of the receiver element together with the associated parameters according to the DOAS (Differential Optical Absorption Spectroscopy) method.

Fig. 2

Absorption spectra from sample gases (schematic)



The method is based on the physical capability of gas molecules to absorb light energy in the typical wavelength ranges for the respective types of gas. Spectral dissection in the receiver element captures the absorption of gas molecules at characteristic points of the spectrum in the IR wavelength range 200 to 400 nm. Corresponding sections within this range are evaluated depending on the device type. The optimized algorithms used here ensure concentrations of gas components searched for are determined without cross-sensitivity to other gases. To do this, the respective evaluation IR range for the components to be measured is defined so that no “foreign gases” disturb the spectrum structure.

2.3 Zero adjust (description)

Zero adjust basically fulfills two tasks:

- Adjustment of the device in the measuring path
- Determination of reference values for check cycles
- Zero adjust is started manually (→ p. 66, §5.1)

2.4 Reference cycle (description)

The reference cycle serves to compensate changes in lamp intensity (e.g. through aging) and contamination effects in the SR-unit.

The measured concentration values are retained (e.g. on analog outputs) during the reference cycle.

The reference cycle interval is adjustable (→ p. 34, §4.2.3.1).

2.4.1 Reference cycle sequence

- 1 Determining the spectrometer offset
- 2 Setting the lamp parameters
- 3 Determining the aperture signal (dark signal)
- 4 Determining the reference spectrum
- 5 Determining the 4Q offset

2.5 Check cycle (description)

The check cycle serves to check the zero and reference point for each component without using test gases.

The check cycle meets the requirements of EN14181 and obviates “drift monitoring with test gases” according to QAL3.

The check cycle interval is adjustable (→ p. 34, §4.2.3.1).

2.5.1 Control elements

Control elements that can be swiveled in are used for measuring during the check:

- 2 grating filters:
 - Grating filter 1: Finely woven grating with 51% transmission and 0.3 absorption
 - Grating filter 2: Finely woven grating with 6.3% transmission and 1.2 absorption

- NO cell:
 - Active length 5 mm
 - Filling 20% NO in N₂
 - Single beam path

Relative to the double beam path used in measuring operation, this gives a C x L of approx. 500 ppm x m or 600 mg/m³ x m

- 1 zero point reflector

2.5.2 Check cycle (function)

The following influencing factors are the main cause of drifts in concentration measurement:

- Wavelength scale drift
- Spectral resolution drift
- Absorption drift

Methods to monitor these influencing factors are described in the following.

2.5.2.1 Wavelength scale monitoring

The grating spectrometer used in the GM32 shows, due to the design, wavelength scale drifts caused by mechanical effects (e.g. length changes caused by temperature fluctuations). Compensation is realized through continuous determination of this drift during measuring operation and calculation back to the original state through numeric interpolation.

Wavelength drift is monitored in two stages during the check cycle:

- 1 Determining the position of an NO absorption line ensures correct functioning of wavelength drift compensation.
To do this, a spectrum is examined with the NO cell swiveled in and the line position deviation calculated as compared against the original device state.
A maintenance request is signaled when this deviation exceeds a limit value (normally 0.046 nm)
- 2 The overall wavelength drift is checked relative to the original device state. A maintenance request is also signaled here when a limit value (normally 1 nm) is exceeded.

These measures ensure wavelength drift is compensated correctly and therefore has no influence on measured values.

2.5.2.2 Resolution drift monitoring

Spectral resolution of the grating spectrometer used can change during operation due to mechanical changes or contamination (scattered light). Resolution deterioration normally leads to a reduction in sensitivity and the device displays values that are too low.

This is why the spectral resolution is examined during the check cycle.

To do this, the half-width of an NO absorption line is calculated from the NO cell spectrum (see above).

A maintenance request is signaled when this deviation exceeds a limit value (normally 0.035 nm) as compared against the original state.

This check ensures resolution changes are detected *before* device sensitivity changes significantly.

2.5.2.3 Absorption drift monitoring

Measured values output by the device have, to a large extent, a linear association with the measured absorption. The consequence is that errors during absorption measurement have a direct effect on measured values. Possible error sources for absorption determination are fluctuating scattered light intensities in the device as well as erroneous determination of the dark current of the detector array.

Two grating filters that serve as standard are used to check absorption determination. The absorption for each of the two filters is determined during the check cycle and compared against the values in original state. The deviations for both filters are averaged and this value checked against a limit value (normally 2% of the measured value). A maintenance request is signaled when the limit value is exceeded.

2.5.2.4 **Determining the zero and reference point**

The zero point is determined by creating a zero spectrum by swiveling in a zero point reflector during the check cycle. This spectrum corresponds to measurement with a measuring path free from gas.

Based on this zero spectrum and using the device calibration function, the associated measured concentration values for individual components are determined and output.

A maintenance request is signaled when one of these zero values exceeds a limit value (normally 2% of the FS).

The absorption measurement deviation determined above is used to calculate the reference value. The value

$$rp = (70\% + [\text{Absorption deviation in \%}]) \times FS$$

is output as reference value for all components.

As described above, there are three causes for reference point drift:

- Wavelength scale drift
- Resolution drift
- Absorption drift

The measures described in → §2.5.2.1 and → §2.5.2.2 keep the first two drift causes within very strict limits. This is why the value *rp* gained from absorption measurement is representative for reference point drift for all measuring components.

2.5.2.5 **Monitoring the cell NO filling**

The check cycle only runs correctly when the NO cell filling is sufficient.

Determining a value for an NO absorption line serves to monitor the cell concentration.

A maintenance request is signaled when the value sinks below 0.1 absorption units.

2.5.2.6

Summary

Table 1

Evaluation of data determined during the check cycle

	Purpose	Data used	Evaluation	Resulting action
1	Check wave-length scale overall drift	Current wave-length scale	Determine shift of wavelength scale against scaling	Maintenance request signal when shift above limit
2	Check wave-length drift compensation	NO spectrum	Determine NO peak position	Maintenance request signal when deviation from value against scaling over limit
3	Check resolution	NO spectrum	Determine NO peak width	Maintenance request signal when deviation from value against scaling over limit
4	Check absorption scale	Grating filter spectra	Grating filter absorption change relative to scaling	Control value ¹ output for individual components. Maintenance request signal when deviation from nominal value over limit
5	Check zero point	Zero spectrum	Evaluate zero spectrum with calibration function	Zero value output for individual components Maintenance request signal when deviation from zero over limit
6	Check cell filling	NO spectrum	Determine NO peak height	Maintenance request signal when peak height under 0.1 absorption units

¹ Control value calculated as (70%+ grating absorption deviation in %) x FS

2.5.3

Events and signals output

2.5.3.1

Check cycle signals

The *Not_measuring* signal is active during the check cycle.

This signal can be sent via a digital output or read out via the OPC interface.

2.5.3.2

Maintenance request signal

The *Maintenance_request* signal is activated when an unallowed drift is determined or the NO cell is empty → p. 13, §2.5.2) and, at the same time, an appropriate message entered in the Logbook. The *Maintenance_request* signal can be sent via a digital output or read out via the OPC interface.

2.5.3.3

Zero and reference point level output

The zero and reference point values determined can be output via the analog outputs. This is done - depending on the parameter settings - directly after the check cycle or on demand (via a digital input). The *Output_control_values* signal is active during output and can also be sent via a digital output. The zero values are first output for 90 s and then the reference values for 90 s. The *Not_measuring* signal is *not* active during the output.

As an alternative, the zero and reference values of the last check cycle can be read out using SOPAS ET menu /*Diagnosis/Check cycle*.

2.6 Contamination compensation

During the reference cycle, the zero point reflector serves to compensate contamination occurring in the SR-unit. Contamination of the front window or measurement reflector cannot be compensated in this manner. Light absorption also occurs during measuring operation through dust in the flue gas. Both of these effects cannot be determined separately due to the principle used and are compensated together using a basic line correction during spectrum evaluation (→ p. 11, §2.2).

2.6.1 Monitoring contamination

As described above, contamination outside the SR-unit cannot be differentiated from dust effects. However, both effects are monitored together by continuous control of the maximum absorption measured in the spectral range of the evaluation. A *maintenance request* is signaled when this absorption rises above a warning value that can be set as a parameter (standard: absorption 1.8). *Failure* is signaled when a limit value is exceeded (standard absorption 2.0).

2.7 Automatic adjustment with alignment

Swivel movements of the reflector relative to the SR-unit can occur during measuring operation. This can be caused, for example, when the probe bends due to varying gas temperatures. These swivel movements are compensated continually on the GM32 by the built-in tracking mirror. Motors tilt this mirror in two directions and can thereby compensate occurring swivel movements.

2.8 Measuring probe in detail

Fig. 3 GPP measuring probe with ceramic filter and ceramic/teflon filter



Integrated sensors

All probe versions have a built-in pressure sensor as well as an integrated temperature sensor PT 1000 that continually measures the medium temperature of the probe in the active measuring path.

The measured data are transmitted via the CAN bus interface of the measuring probe.

EPA conformity

When using a GPP probe/EPA probe, an audit measurement conforming to EPA CFR 40 Part 60 and/or Part 75 can be carried out with the device fitted.


GMP probe with open measuring gap

Shortest reaction times and high temperature stability characterize the GMP series measuring probes. Continuous purge air feed is required for operation. The air outlet runs in the duct offset 90° to the gas flow (Directed Purge Air).

The GMP probe has a locking device on the opening for sample gas that is activated with a lever on the probe flange.

Fig. 4

GMP measuring probe (with open measuring gap)




GPP – gas diffusion probe in dry or wet version

This version is more suitable for higher dust contents because, on GPP probes, dust particles are separated on the filter element and therefore kept away from the measuring path. To allow an EPA-conform audit measurement as well as with low flow speeds or irregular flow profiles, the GPP (**G**as **P**ermeable **P**robe) should also be chosen.

Fig. 5

Measuring gap of the GPP measuring probe



Both variants of the GPP differ in the respective filter which provides suitability for different application areas.

They are fitted with an automatically controlled heater to reliably prevent condensate on the optical interfaces.

The heating control is located between the assembly and device flange in a housing connected firmly with the probe. The electronics for heating control and temperature and pressure measurement are protected safely in a stable cast housing that forms the section of the measuring probe between the duct flange and SR-unit. Both the electrical connections for CAN bus and power supply as well as the test gas connection and pressure measurement connection are fitted on this housing and serve manual activation of the audit measurement according to EPA Guideline CFR 40, Part 60 or Part 75.

GM32

3 CAN Connection

CAN connection
SCU connection


3.1 GM32 CAN connection

The CAN bus is a 2-wire bus system to which all bus participants are connected parallel (i.e. with short stub lines).

- Each end of the CAN bus must be closed off with a $120 \pm 10\%$ Ohm terminating resistor (prevents reflections).
This is also required for very short line lengths.

Fig. 6

CAN bus principle



Terminating resistors must be activated for first and last bus participants.

The terminating resistor must be deactivated for middle bus participants.

- ▶ GM32 default value:
 - Probe: Resistor activated (last participant of a stub line)
 - Gateway: Resistor activated (last participant).
The LED in the gateway is on.
- ▶ Activating or deactivating terminating resistors → Operating Instructions of bus participants.
 - ▶ For GM32: → p. 24, §3.3 and → Operating Instructions “Modular System I/O”.

Stub lines lead to reflections on the bus, therefore:

- ▶ Avoid stub lines whenever possible, otherwise limit these to max. 10 m.

CAN wiring:

- ▶ Maximum length of the CAN bus: 1000 m
- ▶ Cable twisted in pairs and shielded
 - Surge impedance: 120 Ohm
 - Capacitance: ≤ 60 pF/m.

Connect the shield across the complete bus and only ground galvanically at one location (prevents ground loops).

3.2

SCU connection (overview)

As an option, the GM32 (and further analyzers) can be operated using an SCU (System Control Unit).

The SCU is an operating unit for comfortable, powerful analyzer control.

The following actions can be performed on the analyzers via the SCU:

- Control, parameter setting and display.
- Process and save measured values.
- Remote diagnosis.

The SCU is connected with the analyzers via a system bus.


The SCU is operated via a touchscreen or a PC.



Further information on the SCU → “SCU” Operating Instructions

Fig. 7

SCU concept




3.3

Connecting the GM32 to the CAN bus and to the SCU

Connection of the GM32 to the CAN bus, here as example to the SCU.

Fig. 8

Example: CAN connection of the SCU to the GM32



● Terminal assignment:

CAN designation	Terminal in GM32 connection unit	Pin on CAN1 of SCU
CAN H	5	7
CAN L	6	2
gnd	7	3
Shield on PE		

3.3.1 **Settings in GM32**

Menu: Parameter/Device Parameter (→ p. 34, §4.2.3.1)

Valid for the CAN interface to SCU.

- Address for SCU interface: 17 (preset for 1 device)
(range available for further devices: 16, 18, .. 31)
- Baudrate: 125 (preset)

3.3.2 **Settings in the SCU**

- Set in the SCU (→ Operating Instructions of the SCU):
 - Baudrate: Corresponding to the GM32 setting
 - CAN: 1 (always)

3.4

Mapping Table**Measured values on SCU – Measured value (MV)**

Index	Measured value				
MV01..MV16	Identifiers are defined during calibration. Maximum 14 gases + temperature and pressure. Physical unit: mg/m ³ operation				
Version	MV01	MV02	MV03	MV04	MV05
All					
GM32-1	SO ₂				
GM32-2	SO ₂	NO	NO _x		
GM32-3	SO ₂	NO	NO ₂	NO _x	
GM32-4	NO	NO _x			
GM32-5	SO ₂	NO	NH ₃	NO _x	
GM32-6	NO	NO ₂	NH ₃	NO _x	
GM32-7	NO	NO ₂	NO _x		
GM32-8	NO	NO ₂	NO _x		
GM32-9	SO ₂	NO	NO ₂	NH ₃	NO _x
TRS ¹	H ₂ S				
	TRS				
	H ₂ S	SO ₂	NO		
	TRS	SO ₂	NO		
	H ₂ S	SO ₂	NO	NH ₃	
	TRS	SO ₂	NO	NH ₃	
	TRS	H ₂ S	MMK, DMDS, DMS, SO ₂ , NO, NH ₃		

¹ Measured value assignment depending on configuration

Control values on SCU – Monitor values (MO)

Index	Control value
M001..M016	Monitor value 1 = zero from MV01 Monitor value 2 = range from MV01 Monitor value 3 = zero from MV02 Monitor value 4 = range from MV02 Monitor value 27 = zero from MV14 Monitor value 28 = range from MV14

Help values from SCU to GM32 - Help value (HV)

Index	Reference value
HV01	Temperature [K]
HV02	Pressure [hPa]

GM32 operating mode – State (S)

Index	Operating mode
S01	Initialization
S02	Measuring
S03	Maintenance
S04	RCycle
S05	CCycle
S06	ZeroAdjust
S07	Alignment
S08	Boxmeasuring
S09	Restart

GM32 status (F, M, U, C, E)

Index	Diagnostic message
F01..F64	Failure
M01..M32	Maintenance
U01..U08	Uncertain
C01..C08	Check
E01..E16	Extended

Status of measured values (MV)

Index ¹	Diagnostic message
MVxxF01..F64	Failure
MVxxM01..M08	Maintenance
MVxxU01..U16	Uncertain
MVxxC01..C08	Check
MVxxE01..E32	Extended

¹ xx = 1..16

GM32

4 Operation (Specialist Menus)

SOPAS ET
Menus

4.1

SOPAS ET (description)

Operator menus and measured value displays are also available on an external PC via Ethernet for user comfort (with the engineering tool SOPAS ET).

The **SICK Open Portal for Applications and Systems** (SOPAS) is an engineering tool for communication with analyzers and sensors.

SOPAS is based on the following techniques:

- Device communication via Ethernet (TCP/IP).
- Common engineering tool for various product lines.
- Universal device description file as data source for all relevant device data and parameters required for communication and display.
- Interfaces for direct communication with analyzers or SCU.

Methods for accessing SOPAS-based analyzers

- Via the SCU operating unit (→ p. 23, §3.2).
- Via a PC with the program (Engineering Tool) SOPAS ET




The menu structure and representation of menus are principally identical on the SCU and on the PC with SOPAS ET. The representation on the SCU is adapted to the smaller monitor.



For more information on the SOPAS concept, see the Help menu of SOPAS ET.

Fig. 9

SOPAS system concept



4.2

Menu tree (in SOPAS ET)

GM32		Reference	Level ¹
— ----	 Measured values	→ p. 33, §4.2.2	A
	--  Bargraph measured values	→ p. 33, §4.2.2.1	A
— ----	 Parameters	→ p. 34, §4.2.3	A
	--  Device parameters	→ p. 34, §4.2.3.1	A
	--  Spectrometer	→ p. 35, §4.2.3.2	S
	--  Spectral analysis	---2	S
	--  Coefficients	→ p. 35, §4.2.3.3	A
	--  Coefficients boxmeasuring	→ p. 36, §4.2.3.4	S
	--  Check cycle	---2	S
	--  Adjustment sensor	→ p. 36, §4.2.3.5	S
	--  Mirror tracking	---2	S
	--  Data storage	---2	S
	--  Logbook configuration	→ p. 37, §4.2.3.6	S
	--  Analog output	→ p. 38, §4.2.3.7	A
	--  Analog input	→ p. 38, §4.2.3.8	A
	--  Digital input	→ p. 39, §4.2.3.9	A
	--  Digital output	→ p. 39, §4.2.3.10	A
	--  Hardware Map	→ p. 41, §4.2.3.11	S
	--  Options	---2	S
— ----	 Adjustment	→ p. 42, §4.2.4	A
	--  Alignment	→ p. 42, §4.2.4.1	A
	--  Zero adjust	→ p. 43, §4.2.4.2	S
	--  Boxmeasuring	→ p. 44, §4.2.4.3	A
— ----	 Diagnosis	→ p. 45, §4.2.5	A
	--  Device information	→ p. 45, §4.2.5.1	A
	--  Software versions	---2	S
	--  Logbook	→ p. 45, §4.2.5.2	A
	--  Check cycle	→ p. 47, §4.2.5.3	A
	--  Check values Service	---2	S
	--  Sensor values	→ p. 47, §4.2.5.4	A

¹ A = user level "Authorized user".

S = user level "Service".

² Not described in this Manual. Please contact SICK Customer Service.

— ----		Maintenance	→ p. 56, §4.2.6	A
	--	 Operating mode switch	→ p. 56, §4.2.6.1	A
	--	 Lamp hour meter	→ p. 56, §4.2.6.2	A
	--	 Save EEPROM	→ p. 57, §4.2.6.3	A
	--	 Load EEPROM	→ p. 57, §4.2.6.4	A
	--	 Save software debug	---2	S
	--	 Reset parameter	---2	S
	--	 Service log	---2	S
		 Maintenance Functions	---2	S
— ----		Start-up Assistant	→ p. 57, §4.2.7	S

4.2.1

Changing the user level

The menu tree shown above is the structure for user level “Authorized user” and “Service”.

- Enter a password for the corresponding user level:

User level	Password
Authorized client	HIDE ¹
Service	GM32SERVICE

¹ Upper case mandatory


4.2.2 Measured values

4.2.2.1 Bargraph measured values

Menu: GM32/Measured values/Bargraph measured values

This menu shows:

- Measured values
The measured value display can be configured. It is defined by the assignment of analog inputs.
- Operating mode of the analyzer
- Operating mode of the components
- Current sample gas pressure
- Current sample gas temperature



4.2.3 Parameters

4.2.3.1 Device parameters

Menu: GM32/Parameter/Device parameter

- Check cycle component: Checkmark: Check cycle component installed.
- I/O module configuration: Checkmark: I/O module installed.
- Terminating resistor (for the CAN bus to the I/O modules):
 - Checkmark set: Terminating resistor will be fitted (physically).
 - Checkmark not set: No terminating resistor fitted.
- Process optic :
 - No device: Probe installed without CAN bus
 - Probe: Probe installed with CAN bus
 - CD: Cross Duct version
- Address for SCU interface
- Baud rate
- Terminating resistor (for CAN interface to SCU):
 - Checkmark set: Terminating resistor will be fitted (physically).
 - Checkmark not set: No terminating resistor fitted.
- Location designation.
- IP-Configuration (checkmark): IP address change allowed / not allowed.
- Address for SCU interface (only when SCU used): Default: 17.
- Measuring distance: Active measuring path. The measuring path is read in from the probe via the CAN bus).
- Measuring cycle interval: Default: 5 s.
- Reference cycle interval: Interval for the reference cycle.
Default 60 min. (change only possible for user level “Service”).
- Check cycle interval: Interval for the check cycle.
0 min. means: Switched off (change only possible for user level “Service”).

- Temperature optic: Default 323 (change only possible for user level “Service”).
- Threshold alignment: No alignment below this value (editing only possible for user level “Service”).
- Set time: Clicking “Set time” triggers a system reset.
Reconnect to analyzer again afterwards.

All parameters can only be set with user level “Service” (exception: Time).

4.2.3.2

Spectrometer

Menu: GM32/Parameter/Spectrometer

Spectrometer			
Serial number spectrometer	CMOS_0499		
Averaging	5		
Initial current for lamp	120	mA	
Minimum exposure	100	ms	Maximum exposure 500 ms
Maximum lamp intensity (zeropoint reflector)	10000		
Lambda_c0_act	1.994494E2	New value (Switch device to maintenance first!)	0E0 Save
Lambda_c1_act	2.1689171E-1		
Lambda_c2_act	2.8013719E-7		
Lambda_c3_act	-5.730127E-9		
Spectral resolution	0.476		
Spectrometer temperature	318	K	
Peltier ☒ Set temperature for peltier	318	K	
Start diode for analysis	50	End diode for analysis	1000

- Internal parameters of the spectrometer (change only possible with user level “Service”).

4.2.3.3

Coefficients

Menu: GM32/Parameter/Coefficients

Coefficients ($y = C2 \cdot x^2 + C1 \cdot x + C0$)				
Measured value 01	SO2	C2	0E0	C1 1E0 C0 0E0
Measured value 02	NO	C2	0E0	C1 1E0 C0 0E0
Measured value 03	NO2	C2	0E0	C1 1E0 C0 0E0
Measured value 04	NH3_QE	C2	0E0	C1 1E0 C0 0E0

- Polynomial to adapt measured values.

4.2.3.4

Coefficients boxmeasuring

Menu: GM32/Parameter/Coefficients boxmeasuring

Coefficients boxmeasuring ($y = C2 \cdot x^2 + C1 \cdot x + C0$)

Measured value 01	SO2	C2	0E0	C1	1E0	C0	0E0
Measured value 02	NO	C2	0E0	C1	1E0	C0	0E0
Measured value 03	NO2	C2	0E0	C1	1E0	C0	0E0
Measured value 04	NH3_QE	C2	0E0	C1	1E0	C0	0E0

- Polynomial to adapt measured values to the filter box (change only possible with user level "Service").

4.2.3.5

Adjustment sensor

Menu: GM32/Parameter/Adjustment sensor

Adjustment sensor (lower limit < sum < upper limit)

Upper limit value	1700
Sum of 4Q sensor	1000
Lower limit value	400
Dead time	10 Meascycle

- Internal parameters of sensor alignment (change only possible with user level "Service").

4.2.3.6

Logbook

Menu: GM32/Parameter/Logbook configuration

- This menu serves to set the logbook parameters (change only possible with user level "Service").



Changing the setting deletes the logbook contents.

Input line	Remark
uncompressed	When an error occurs:
- Compressed	- An error counter is incremented.
- Uncompressed	- The error message is saved.
Ring buffer mode	When the logbook buffer is full:
- Warning mode	- The message "Logbook error" is output and no further entries are saved.
- Ring buffer mode	- The oldest entries are overwritten with the current entries. No relevant message is displayed.
Number of logbook messages	Number of logbook entries

4.2.3.7 Analog outputs

Menu: GM32/Parameter/Analog output / Display

Parameter settings for the analog outputs are also used to scale the measured value display (on the operator panel and in SOPAS ET).

Analog output / Display

4 mA ☐ Output of control values ☒ Hysteresis 10 %

Channel	Gas	Source	Unit	MR1 Start	MR1 End	MR2 Start	MR2 End	Control value	Signal
1:	SO2	MV01	g/m3	0	2	0	100	☒	☒
2:	NO	MV02	mg/m3	0	30	0	100	☒	☐
3:	NO2	MV03	mg/m3	0	50	0	100	☒	☐
4:	NH3	MV04	mg/m3	0	20	0	100	☒	☐

- Electrical zero point “Live Zero”: 0 or 4 mA
- Output of control values after check cycle. Checkmark: Yes
Zero and control values are output after the check cycle:
 - First of all the zero value for 90 seconds
 - Then the control value for 90 seconds.
- Hysteresis: Threshold for switching from measuring range 1 to measuring range 2, or vice-versa.
- Source: The signal source is defined during the factory calibration.
- Physical unit
- Start/End: Calibrated measuring range 1 or measuring range 2.
- Signal: Digital output on which measuring range 2 is signaled.
- Control value output can be activated or deactivated individually for each of the 16 analog outputs.
Control value: Checkmark means: Output the control value for this channel.

4.2.3.8 Analog inputs

Menu: GM32/Parameter/Analog input

Analog input

Channel	Source	Unit	Start	End
1:	T	K	0	100
2:	T	K	0	100

- Signal sources for measured variables temperature T / pressure P)
- Physical unit of the measured variable
- Electrical zero point “Live Zero”: 0 or 4 mA
- Analog input range for signal source: Start / End

4.2.3.9

Digital inputs

Menu: GM32/Parameter/Digital input

Source	Action
Check_cycle	Triggers operating mode switch to "Check cycle"
Disable_check_cycle	Suppresses the check cycle
Maintenance	Triggers operating mode switch to "Maintenance"
No_function	No function assigned
Output_control_values	Outputs control values
Purge_air	Purge air monitoring
Inverted	Inverts the signal

4.2.3.10

Digital outputs

Menu: GM32/Parameter/Digital output

Signal	Cause
Failure	A malfunction has occurred
Maintenance_Request	A maintenance request is available
Uncertain	The GM32 is in an indefinite state
Not_measuring	Zero adjust is running Check cycle is running Operating mode "Maintenance" Operating mode "Boxmeasuring" (filter box measurement)

Signal	Cause
Extended	A message is available
Output_control_values	Outputs control values
Measuring_range	Outside measuring range
Purge air failure	Error in purge air supply
No_function	No function assigned
Inverted	Inverts the signal

4.2.3.11 Hardware Map

Menu: GM32/Parameter/Hardware Map

This menu serves to enter the I/O modules in the sequence (left to right) in which they have been installed in the connection unit (→ connecting diagram).

Table 2 Module Table

Module	Description
A002	2 analog outputs
AI02	2 analog inputs
DO04	4 digital outputs
DI04	4 digital inputs

- 1 Enter modules.
Line 1: Module 1
Line 2: Module 2
etc.
- 2 Use “Edit” to enter installed modules in the corresponding line.
Example: The first module is an “AO02” module.
- 3 “Save”.
- 4 Store the modules with “Restart IOProxy”.

4.2.4 Adjustment

4.2.4.1 Alignment

Menu: GM32/Adjustment/Alignment

This menu serves to align the optics system.

The L adjustment of the device flange is used for optical alignment.



For further information on optical alignment → GM32 Operating Instructions

Alignment

Operation ● Amplifier adjustment ○

Operating mode: Measuring Start alignment

Offset A: 1071 Offset B: 1065

Offset C: 1099 Offset D: 1088

A: 230 4Q A: 230.0 0.0 1024.0

B: 201 4Q B: 204.0 0.0 1024.0

C: 199 4Q C: 199.0 0.0 1024.0

D: 211 4Q D: 211.0 0.0 1024.0

X: 0.028

Y: 0.045


Measuring

- 1 *“Start alignment”*
Measuring operation is suspended during alignment.
Tolerances:
X: ± 0.05
Y: ± 0.05
- 2 Click *“Measuring”* after completing alignment.

4.2.4.2

Zero adjust (Service level only)

Menu: GM32/Adjustment/Zero adjust



The screenshot shows the 'Zero adjust' menu. At the top, it says 'Zero adjust'. Below that, 'Operation' is followed by a green circle icon. 'Operating mode' is followed by a text box containing 'Measuring'. At the bottom, there are two buttons: 'Start zero adjust' and 'Start measuring mode'.



Zero adjust description → p. 12, §2.3



The GM32 must be at operating temperature.
► Otherwise observe the 2 hour warm-up time.



Zero adjust is relative to the composition of the gas currently in the measuring path.
● Inert gas must be in the active measuring path

Zero adjust takes about 5 minutes.

The GM32 switches to operating mode “Maintenance”.

“Start measuring mode” automatically starts a reference cycle.

The GM32 then switches to operating mode “Measuring”.

4.2.4.3 Boxmeasuring (Service level only)

Menu: GM32/Adjustment/Boxmeasuring

Boxmeasuring

Boxmeasuring ☐

Gas temperature K

Maximum lamp intensity for boxmeasuring

Operating mode

Malfunction ☐ Maintenance Request ☐

0.0	SO2	0	mg/m3	1500.0	Malfunction	☐	Maintenance	☐	Uncertain	☐
0.0	NO	0	mg/m3	1100.0	Malfunction	☐	Maintenance	☐	Uncertain	☐
0.0	NO2	0.0	mg/m3	160.0	Malfunction	☐	Maintenance	☐	Uncertain	☐



Boxmeasuring description → p. 68, § 5.2

4.2.5 Diagnosis

4.2.5.1 Device information

Menu: GM32/Diagnosis/Device information

Device information

Serial number

Calibration package: Test_5K
 Device: GM32
 Components:
 SO2
 NO
 NO2
 NH3
 NOx
 Temperature Range: 20-150°C
 Spectrometer ID: cmos103
 Comment: resolution etvl Cndern
 Packed: 19.05.2010 16:48:14.09

Firmware version

4.2.5.2 Logbook

Menu: GM32/Diagnosis/Logbook

Max. number of entries: 6000.

(Representation: Uncompressed data storage)

Measure

Diagnosis/Logbook

100% Entries: 117

Display all messages

Reset
Update
Last Data
Next Data

No.▼	Device Name	Text	Class	Date Start	Time Start	Date Stop	Time Stop
1	System	Systemstart	X	00/01/01	00:00:13	-----	-----
2	System	Systemstart	X	00/01/01	02:32:07	00/01/01	00:00:12
3	System	Systemstart	X	00/01/01	00:31:23	00/01/01	00:00:09
4	System	Systemstart	X	00/01/01	00:21:13	00/01/01	00:00:12

Designation	Remark
	Fill level of logbook in %. When the characters are <i>red</i> : The logbook is full. Warning mode: Further entries are not accepted. Ring buffer mode: Oldest entries are overwritten.
	Data storage: Symbol <i>not crossed out</i> : Compressed. Symbol <i>crossed out</i> : Uncompressed.
	Significance and default: → p. 37, § 4.2.3.6
	Ring buffer mode
	Warning mode
	Significance and default: → p. 37, § 4.2.3.6
Entries	Number of entries of selected filter.

Designation	Remark
Filter for messages	Only the filtered messages are shown. - Show active failures - Show all failures - Show active maintenance request - Show all maintenance requests - Show active uncertain - Show all uncertain - Show active extended messages - Show all extended messages - Show active messages - Show all messages Classification → Further on in this Table.
Reset	Clear all entries.
Export (Only in SOPAS ET)	All entries selected via the filter (→ Further back in this Table) are saved on the PC as .log file. Format: CSV (comma-separated list). Can be read in EXCEL, for example.
Update	Update display of logbook entries.
Last Data	Scroll back.
Next Data	Scroll forward.
▲▼	Sort in ascending/descending order. To switch sorting on or change sequence: Click column header.
	Current number of message. <i>Red LED</i> : Message still pending. <i>Green LED</i> : Message no longer pending.
Device Name	Activating element: System, measured value identifier (sample gas component), subassembly, Evaluation module
Entries ¹	Number of times errors have occurred. Significance and default value: (→ p. 37, § 4.2.3.6)
Text	Logbook message.
Class	F = Failure M = Maintenance request C = Check U = Uncertain X = Extended message / extended
Date Start	Format: yy-mm-dd For "Uncompressed": Occurrence of message. For "Compressed": Last occurrence of message.
Time Start	Format: hh:mm:ss For "Uncompressed": Occurrence of message. For "Compressed": Last occurrence of message.
Date Stop	Format: yy-mm-dd For "Uncompressed": Clearing of message. For "Compressed": Last clearing of message.
Time Stop	Format: hh:mm:ss For "Uncompressed": Clearing of message. For "Compressed": Last clearing of message.

¹ Only for compressed data storage

4.2.5.3 Check cycle

Menu: GM32/Diagnosis/Check cycle

The check cycle values are the values measured during the last check cycle. “Zero” means no check cycle has been performed yet.



The check values are set to “Zero” during a restart.

Check cycle							
1:	SO2	Zero	0	0 mA	Span	0	0 mA
2:	NO	Zero	0	0 mA	Span	0	0 mA
3:	NO2	Zero	0	0 mA	Span	0	0 mA

- Zero: Zero point as physical and current value
- Span: Control point as physical and current value relative to the upper measuring range value

4.2.5.4 Sensor values

Menu: GM32/Diagnosis/Sensor values

Spectral analysis

Menu: GM32/Diagnosis/Sensor values/Spectral evaluation

Spectral evaluation			
Temperature	300	K	Substitute
Pressure	1013	hPa	Substitute
Measuring distance	500	mm	
Timer visor no signal	0	s	Timer absorption failure
	0	s	
1:	SO2	Qualifier	0.295
2:	NO	Qualifier	0.055
3:	NO2	Qualifier	0.191
4:	NH3	Qualifier	0.951
5:	NOX	Qualifier	0

- Sources and values for spectral evaluation

Spectrometer

Menu: GM32/Diagnosis/Sensor values/Spectrometer

Spectrometer values from last reference cycle.

Spektrometer		
Belichtungszeit	200	ms
Lampenimpuls	64	mA
Spektrometer-Offset	4	
Spektrometertemperatur	318,161	K
Zeilentemperatur	318,161	K
Peltierstrom	557,396	mA

Adjustment sensor

Menu: GM32/Diagnosis/Sensor values/Adjustment sensor

Limit values for lamp pulse and amplification settings are entered in the *Parameter/Adjustment sensor* menu (→ p. 36, § 4.2.3.5).

Adjustment sensor			
Lamp pulse	100	mA	
Amplifier measurement reflector	118		
Offset A	18	Offset B	30
Offset C	8	Offset D	6
A	359	B	355
C	318	D	321
Sum	1291	Counter	0

- Lamp pulse: Lamp pulse current during the 4 quadrant measurement. Changes depending on the total from the 4 quadrants.
Criterion: Amplifier measurement reflector at 255.
- Amplifier measurement reflector: Amplification of the 4 quadrant measurement. Changes depending on the total of the 4 quadrants (max. 255).
- Offset: Offsets are determined during the check cycle or after amplifier switching.
- A .. D, Sum: Current signals.
- Counter: Delay time before correction made.

Mirror tracking

Menu: GM32/Diagnosis/Sensor values/Alignment

Current values from alignment.

Alignment	
Visior fault 	
Steps direction x	0
Deviation direction x	0.031
Number of movements direction x	6
Steps direction y	0
Deviation direction y	0.008
Number of movements direction y	6



A device restart sets the number of movements to zero.

Process optic

Menu: GM32/Diagnosis/Sensor values/Process optic

Current analyzer values

Process optic	
Probe	
Temperature	273.15 °K
Pressure	0 hPa
Measuring distance	0 mm
Tube length	0 mm

4.2.5.5

Temperature unit*Menu: GM32/Diagnosis/Sensor values/Temperature unit*

Current analyzer values

Temperature unit		
Temperature optic	309.023	K
Temperature spectrometer	318.161	K
Temperature detector	318.253	K
Peltier current	464.497	mA
Temperature transistors	320.75	K
Voltage	21.389	V
Current optic heating	0.01	A
Current spectrometer heating	0.077	A
Fan lamp	103	Hz
Fan optic	105	Hz
Fan spectro	299	Hz

4.2.5.6

Spectra

Significance of individual spectra fields:


Field	Significance
Continuous	The displayed spectrum is continuously refreshed.
Time	Time interval of continuous display.
Record	Activates data recording → Enter name and location of file for data recording → Save
Refresh	Refresh spectrum.
Combine values Show individual measured points	Combine measured value points (line display) Show individual measured value points.



Click two points in the spectrum to zoom sections of the spectrum.

Measurement reflector

Menu: GM32/Diagnosis/Sensor values/Spectra/Measurement reflector




- Measurement reflector raw signal
- Maximum measurement reflector signal on the zero path: 45000 .. 50000 digits.
Change with user level Service in menu *Parameter* → *Spectrometer* → *Maximum lamp intensity*.

Zero point reflector

Menu: GM32/Diagnosis/Sensor values/Spectra/Zeropoint reflector

Raw signal measured on the zero point reflector. The nominal value corresponds to the maximum lamp intensity (→ p. 35, §4.2.3.2).




- Zero point reflector signal
- Change with user level Service in menu *Parameter* → *Spectrometer* → *Maximum lamp intensity*.

Aperture signal (dark signal)

Menu: GM32/Diagnosis/Sensor values/Spectra/Aperture signal


Measurement on blanking diaphragm. Only updated during the check cycle.



Absorption

Menu: *GM32/Diagnosis/Sensor values/Spectra/Absorption*

Current absorption spectrum. A message is output when the warning or error threshold is exceeded (→ p. 35, §4.2.3.3). Areas at start and end are masked out (→ p. 35, §4.2.3.2).




Filtered absorption

Menu: *GM32/Diagnosis/Sensor values/Spectra/Filtered absorption*

Averaged absorption spectrum.

Preselected filter, e.g. IRR filter (infinite impulse response filter).


Menu */Parameter/Spectral alignment* serves to select the filter (only possible for user level "Service").



Wavelength corrected absorption

Menu: GM32/Diagnosis/Sensor values/Spectra/Wavelength corrected absorption

Output of wavelength corrected absorption




Reference spectrum

Menu: GM32/Diagnosis/Sensor values/Spectra/Reference

This spectrum serves to calculate the absorption and shows the calculated spectrum on the measurement reflector on the smoke-free zero path.

Reference spectrum = (zero point reflector signal - aperture signal (dark signal)) * scaling constant.




Normalisation

Menu: *GM32/Diagnosis/Sensor values/Spectra/Normalisation*

Graphic representation of normalisation (constants for zero adjust).

Normalisation must be between 0.8 and 5.



4.2.6 Maintenance

4.2.6.1 Operating mode switch

Menu: GM32/Maintenance/Operating mode switch

Operating mode switch

Operation

Maintenance

Measuring

Reference cycle

Reference cycle

Check cycle

Check cycle

System reset

Start system reset (After execution establish new connection!)

Field	Significance
Operating mode	Displays the current operating mode.
Maintenance	Switches to operating mode "Maintenance".
Measuring	Switches to operating mode "Measuring"
Reference cycle	Starts a reference cycle
Check cycle	Starts a check cycle
System reset	Resets the system. Then reconnect SOPAS ET and GM32 again (→ p. 30, §4.1).

4.2.6.2 Lamp hour meter

Menu: GM32/Maintenance/Lamp hour meter

Lamp hour meter

Lamp hour meter

1580

 h

Performance lamp

95

 %

Reset lamp hour meter

4.2.6.3 Save EEPROM

Menu: GM32/Maintenance/Save EEPROM

Field	Significance
Start EEPROM saving	EEPROM contents (and therefore the parameters and adjust data) are stored on the Flashcard <i>Attention:</i> Existing data are overwritten.
LED	Goes on when the storage process finishes.

4.2.6.4 Load EEPROM

Menu: GM32/Maintenance/Load EEPROM

Field	Significance
Start EEPROM load	Loads EEPROM contents from the Flashcard. This loads the parameters and adjust data.
LED	Goes on when the load process finishes.

4.2.7 Start-up Assistant

Menu: GM32/Start-Up Assistant

The Start-up Assistant guides you through commissioning step by step.

Criterion: SOPAS ET as from Version 2.32

4.3

Program/data structure (“Firmware”, “Device description file”)

To load data to the GM32 analyzer:

- The GM32 analyzer can be accessed using the SCU (i.e. connect the laptop to the SCU and not to the electronic unit).



Login to SOPAS ET as “Service”.

To receive the password: Please contact SICK Customer Service.

4.3.1

Sequence of settings and file transfers

Action	Imperative	Reference
Set the IP address of the SCU or analyzer.	- When the PCB console has been renewed. - Direct access to SCU/GM32 analyzer. - Login to a different network.	→ p. 59, § 4.3.2
Load firmware (.smf)	Install or update firmware on the GM32 analyzer / SCU. This step also loads the device description file (.sdd).	→ p. 60, § 4.3.3
For SCU only: Load device description file (.smu)	Load the SCU-specific device description file (representation adapted to the SCU display) to the SCU operator panel.	Can be loaded via SOPAS ET.
Load parameters (.sdv)	Load the parameters (device data) (stored on a data medium) in the GM32 analyzer/SCU.	→ p. 63, § 4.3.5.3

4.3.2

Setting the GM32 analyzer/SCU IP address



- ▶ Whenever possible, do not connect your own laptop to the customer network.
It is better to use a customer's laptop (with SOPAS ET).



- If you have to change the GM32 analyzer/SCU IP address temporarily:
- ▶ Note the existing settings.
 - ▶ Reset the old address again after completing the work.


Procedure

- 1 Start SOPAS ET.
 - 2 Network scan assistant.
 - 3 "Network Configuration".
 - 4 "Auto IP configuration" ("Enable AutoIP" must be clicked).
 - 5 "Search".
 - 6 Click the desired device.
If the desired device is not found
 - a) Ping the IP address under DOS:
If no reply is received: Device switched on? Baud rate? Cable crossed (for direct connection)?
 - b) Network scan assistant: Alternately, set *Custom Port 15780* and try again.
 - 7 "Edit" → set the desired IP address.
(Note: SCU and PC must be in the same network segment)
- If you have installed a new PCB console:
- 8 After a new "Network scan", the "Basic Sensor" reports under the set IP address.
 - 9 Drag the "Basic Sensor" into the "Project tree": This loads the "Basic Sensor" device description.
 - 10 Load firmware (→ p. 60, § 4.3.3).

4.3.3 Updating the firmware (GM32 analyzer/SCU)

Fig. 10

Loading the firmware



Existing parameters (device data) are lost when the firmware is loaded.

- First save the existing parameters (→ p. 63, §4.3.5.2) and then load the firmware (→ p. 63, §4.3.5.3). Parameters that no longer match are displayed.

Auxiliary means required

Firmware installation package with .smf file (there are two .smf files for GM32 analyzers)


- Firmware files have the suffix .smf

Procedure

- 1 Connect the PC to the Ethernet interface of the GM32 or SCU.
- 2 Call SOPAS ET.
- 3 “Network scan”.
- 4 Drag the desired device (for an empty PCB console: “Basic Sensor”) into the “Project tree”.
- 5 “Communication”
- 6 “Software Update”
- 7 “Firmware”
- 8 Follow the instructions (load .smf file)
The device description file (.sdd) is loaded automatically at the same time.
 - A second device description file (.smu) must be loaded for the SCU (→ p. 61, §4.3.4).
- 9 “Auto detect”
- 10 In the project tree:
 - a) “Maintenance”
 - b) “Hardware Reset”

4.3.4 Loading the device description file (“Make available to SOPAS ET”)

Fig. 11 Loading the device description file



- The device description files
 - Are located in the SOPAS ET directory (PC dependent).
 - Have the suffixes:
 - GM32 analyzer and SCU: .sdd
 - SCU: .sdd
- The device description file belongs *uniquely* to the firmware version.
- The device description file .sdd is loaded automatically when the firmware version is loaded (→ p. 60, §4.3.3).

Procedure

When the device description files are *not yet* on the GM32 analyzer or SCU:


- 1 Provide the .sdd in any directory.
- 2 Call SOPAS ET.
- 3 “Extras”
- 4 “Module Manager”
- 5 “Install new or update existing module”
- 6 Follow the instructions.

When the device description files are *already* available on the GM32 analyzer or SCU and SOPAS ET is to be made available again:

- 1 Call SOPAS ET.
- 2 “Network scan”.
- 3 In the network scan assistant, devices with invalid or missing device description files are marked: “Load from device”.
- 4 Drag the desired device into the “Project tree”.
This copies the device description file for the selected device into the SOPAS ET directory.

4.3.5 Saving and loading parameters (device data)

Fig. 12 Loading parameters



4.3.5.1 Loading parameters changed in SOPAS ET to the GM32 analyzer/SCU

- 1 Call SOPAS ET.
- 2 Click device (in project tree) with the right mouse button.
- 3 *"Device characteristics"*.
 - a) *"Immediate download"* : Parameters are loaded continuously to the device.
- OR
- b) *"Download on request"*: Load parameters to the device:
 - 1 *"Communication"*
 - 2 *"Switch offline"*
 - 3 *"Download all parameters to device"*
 - 4 *"Switch online"*

4.3.5.2 Saving parameters (GM32 analyzer/SCU) to an external data medium



You can only reset the GM32 to the last usable state after data losses when you have previously saved the data on an external data medium (e.g. a PC).

- Parameter files have the suffix .sdv
- 1 Connect the PC to the Ethernet interface.
- 2 Call SOPAS ET.
- 3 “Network scan”.
- 4 Drag the desired device into the “Project tree”.
- 5 Click the desired device (highlighted blue).



For a complete backup, save both the GM32 analyzer *and* SCU parameters.

- 6 “Project”
- 7 “Export device”
- 8 Follow further instructions.
Recommended file name (e.g. for the SCU): yyyy-mm-dd_scu_customername(.sdv)



Leave one copy of the data with the customer device.
(e.g. on a CD-ROM or on the customer's PC).

4.3.5.3 Loading parameters from an external data medium to the GM32 analyzer/SCU

- Parameter files have the suffix .sdv
- 1 Connect the PC to the Ethernet interface.
- 2 Call SOPAS ET.
- 3 “Edit”
- 4 “Load device data in device”
- 5 Follow further instructions.

4.3.6 Loading data files from the device

The data files contain the results of the check cycles.

- 1 Call SOPAS ET.
- 2 “Communication”
- 3 “File Upload”

GM32

5 Maintenance

Maintenance plan
Maintenance work
Preventative maintenance
Recommended spare parts

5.1 Zero adjust (zero point check)

Special tools and auxiliary means

Auxiliary means required	Part No.	Required for
SOPAS ET		Parameter setting
Ethernet cable (crossover)	6026084	Parameter setting
Mains cable (10m) ¹	2017519	Voltage supply (SR-unit / GPP probe) - Optional when the zero path cannot be reached with the standard cable.
19 mm jaw wrench		

¹ Fit suitable power plug onsite.



Zero adjust is relative to the composition of the gas currently in the measuring path.

- Inert gas must be in the active measuring path
- Starting zero adjust on the operator panel is protected with a password.
Password: 1234

Performing zero adjust is noted in the device Logbook.

The "Not_measuring" signal is active during zero adjust.

5.1.1 Procedure



Assembly and optical adjustment → GM32 Operating Instructions



The GM32 must be at operating temperature.

- ▶ Otherwise observe the 2 hour warm-up time.

- 1 Switch the GM32 to operating mode "Maintenance".
→ SOPAS ET menu *Maintenance* → *Operating mode switch* (→ p. 56, §4.2.6.1)
- 2 Dismount the GM32 from the measuring point.



WARNING: Hazard resulting from gases in the gas duct

Hot and/or noxious gases can escape during work on the gas duct, depending on the plant conditions.

- ▶ Work on the gas duct may only be performed by skilled persons who, based on their technical training and knowledge as well as knowledge of the relevant regulations, can assess the tasks given and recognize the hazards involved.

- 3 Put the GM32 in a safe place and put it into operation with the probe fitted. The probe must only be in inert gas (e.g. ambient air).



If the standard cables are too short:

- Acquire additional mains cables.
- Dismount the power supply unit (→ p. 79, §5.4).

- 4 Clean the window.
- 5 Login as *Service* in SOPAS ET (→ p. 30, §).
- 6 Menu *Adjustment* → *Alignment*.
- 7 Click "*Start alignment*" to move the mirror tracking to the reference position.

- 8 Use the adjusting screws to align the SR-unit (X and Y values between -0.05 ... +0.05).
- 9 Click “Measuring” to terminate the alignment function and to switch the analyzer to measuring operation.
- 10 Call up menu *Diagnosis* → *Sensor values* → *Spectra* → *Measurement reflector* (→ p. 51) to display the raw values from the measurement reflector in digits as graph. The maximum must be between 45000 and 50000 digits.
- 11 Increase or reduce the light intensity for larger deviations. To do this, experiment by adjusting the nominal value for the “Maximum lamp intensity” of the zero point reflector in menu *Parameter* → *Spectrometer* (→ p. 35, §4.2.3.2).
- 12 In menu *Maintenance* → *Operating mode switch* (→ p. 56, §4.2.6.1), start a *Reference cycle* and wait until the analyzer switches back to measuring operation.
- 13 Control the raw values in menu *Diagnosis* → *Sensor values* → *Spectra* → *Measurement reflector* (→ p. 51) often. The maximum value must be between 45000 and 50000 digits. Adapt the maximum lamp intensity further as described in → 11 when the deviation is larger or smaller. When the value on the measurement reflector reaches a value between 45000 and 50000 digits, the GM32 can be compared against zero in the SOPAS ET menu *Adjustment* → *Zero adjust* (→ p. 43, §4.2.4.2).
- 14 Use menu *Diagnosis* → *Sensor values* → *Spectra* → *Absorption* (→ p. 53) to check the smoke-free spectrum. An approximately straight line at “0” should result.
- 15 Use menu *Measured value* → *Bargraph measured value* (→ p. 33, §4.2.2.1) to check the current measured values. The measured values should be close to the zero point and may not deviate from the upper measuring range value by more than 2%.
- 16 Switch the GM32 to operating mode “Maintenance”. Menu *Maintenance* → *Operating mode switch* (→ p. 56, §4.2.6.1)
- 17 Fit the GM32 back onto the measuring point.

5.2

Boxmeasuring**Special tools and auxiliary equipment**

Auxiliary means required	Part No.	Required for
Filter box and accessories		Linearity measurement
Test gases		→ p. 71, §5.2.2
Computer with SOPAS ET		Parameter setting
Ethernet cable (crossover)	6026084	Parameter setting

The gas filter box serves to check the measuring channels with test gases.

**WARNING: Hazards through toxic or combustible gases.**

- The user must be familiar with the hazards through toxic and combustible gases.

5.2.1

Preparing the filter box

- 1 Switch the filter box and pump on.




Warm-up phase for filter box to 80 °C (nominal temperature): Approx. 2 h



WARNING: Risk of breakage for windows in the sample gas chambers

- Operate the filter box with max. 1 bar.

- 2 Set the nominal temperature to 80 °C (“Set” button and arrow buttons).
- 3 Set all valves to the “Zero gas” position.
- 4 Set “Purge valve” to *Purging*.



Temperature displays, preheater and measuring




Recommendation: Use nitrogen (N₂) as zero gas.

5 Fit the drain hose screw fitting to the gas outlets

Fig. 13

Drain hose screw fitting



6 Connect the drain hose to the screw fitting



- Lay the drain hose away from the measuring location (e.g. outdoors)
- ⊗ Do not bend the hose

- 7 Select the suitable Suprasil reflector matching the measuring path (flange-flange) from the filter box accessories.
 - Measuring path <5m (R=-1982)
 - Measuring path >5m (R=-5000).
- 8 Fit the reflector on the filter box.
- 9 Fit the suitable GM32 adapter plate to the filter box

5.2.2

Determining the necessary test gas concentration**1 Calculation of the maximum nominal concentration**SOPAS ET menu *Adjustment* → *Boxmeasuring* (→ p. 44, §4.2.4.3)

Boxmeasuring

Boxmeasuring ☐

Gas temperature K

Maximum lamp intensity for boxmeasuring

Operating mode

Malfunction ☐ Maintenance Request ☒

0.0 SO2 0 mg/m3 1500.0 Malfunction ☐ Maintenance ☐ Uncertain ☐

0.0 NO 0 mg/m3 1100.0 Malfunction ☐ Maintenance ☐ Uncertain ☐

0.0 NO2 0.0 mg/m3 160.0 Malfunction ☐ Maintenance ☐ Uncertain ☐

$$C_{TG} = \frac{FS}{0,15} \cdot \frac{(273 + T)}{273} \cdot \frac{1}{F}$$

 C_{TG} Result is in vppm to order the test gasesFS Measuring range in mg/m³ (see Fig. above)

T Temperature set for the filter box in °C

F Conversion factor ppm to mg/Nm³:NO = 1.34; NO₂ = 2.05; NH₃ = 0.76; SO₂ = 2.86

2 Calculation of nominal values for the various chamber lengths

$$C_{\text{Nom}} = C_{\text{TG}} \cdot F \cdot \frac{273}{(273 + T)} \cdot \frac{\text{AirPressure}}{1013} \cdot L \cdot 0,001$$

C_{Nom} Nominal concentration in mg/m³

C_{TG} Test gas concentration in ppm

T Temperature set for the filter box in °C

F Conversion factor ppm to mg/Nm³:
NO = 1.34; NO₂ = 2.05; NH₃ = 0.76; SO₂ = 2.86

L Chamber length in mm

Table 3 Test gas concentration, example Table

Chamber length [mm]	Nominal [mg/m ³]	Actual [mg/m ³]
25		
50		
75		
100		
125		
150		

5.2.3

Procedure

- 1 Login as *Service* in SOPAS ET.
- 2 SOPAS ET menu *Maintenance* → *Operating mode switch* (→ p. 56, §4.2.6.1).
- 3 Click "*Maintenance*" to switch the GM32 to operating mode *Maintenance*.
- 4 Open the analyzer by loosening the tension clips, remove the hinge pin and take the analyzer off the device flange by loosening the tension clips.
- 5 Mount the analyzer on the filter box and fasten with the tension clips.



► Provide an approx. 20 cm high support for the filter box.

Fig. 14

GM32 with filter box



- 6 SOPAS ET menu *Adjustment* → *Boxmeasuring* (→ p. 44, §4.2.4.3)

- 7 Enter the filter box temperature under “Gas temperature”



Note: Check the nominal value on the filter box.
Standard value: 80 °C (353 K).

- 8 Flush the filter box with zero gas (→ p. 69, §5.2.1)



Recommendation: Use nitrogen (N₂) as zero gas.

- 9 Click “Start boxmeasuring” to start filter box measurement.

Boxmeasuring 1 - 8

Operating mode: Maintenance Start boxmeasuring Maintenance

Unit of measured values: mg/m³ Analog output active: ☒ Gas temperature: 353 K Gas pressure: 1013 hPa

Maximum lamp intensity for boxmeasuring: 15000 Measuring distance: 1000 mm

Diagram time scale: 30 min

<< Back Forward >> Forward >|



- Status line on operator panel: “Boxmeasuring”
- Maintenance request (yellow LED) active

- 10 Wait (up to approx. 20 minutes) until the message “Boxmeasuring (fill in gas!)” appears in “Operating mode”.

Boxmeasuring 1 - 8

Operating mode: Boxmeasuring (fill in gas!) Start boxmeasuring Maintenance


Unit of measured values: mg/m³ Analog output active: ☒ Gas temperature: 353 K Gas pressure: 1013 hPa

Maximum lamp intensity for boxmeasuring: 15000 Measuring distance: 1000 mm

Diagram time scale: 30 min

<< Back Forward >> Forward >|

- 11 SOPAS ET menu *Diagnosis* → *Sensor values* → *Spectra* → *Measurement reflector* (→ p. 51)
The maximum must be between 30000 and 50000.



If this is not the case:

- Experiment with the value for the “Maximum lamp intensity for boxmeasuring”. Click “Maintenance” after changing the value. Change the lamp intensity (maximum 50000) for the filter box accordingly and repeat “Start boxmeasuring”.

Boxmeasuring 1 - 8

Operating mode: Boxmeasuring (fill in gas!) [Start boxmeasuring] [Maintenance]

Unit of measured values: mg/m3 Analog output active: ☒ Gas temperature: 353 K Gas pressure: 1013 hPa

Maximum lamp intensity for boxmeasuring: 50000 Measuring distance: 1000 mm

Diagram time scale: 30 min

<< Back Forward >> Forward >|

Boxmeasuring 1 - 8

Operating mode: Boxmeasuring (fill in gas!) [Start boxmeasuring] [Maintenance]

Unit of measured values: mg/m3 Analog output active: ☒ Gas temperature: 353 K Gas pressure: 1013 hPa

Maximum lamp intensity for boxmeasuring: 50000 Measuring distance: 1000 mm

Diagram time scale: 30 min

<< Back Forward >> Forward >|

- 12 SOPAS ET menu *Diagnosis* → *Sensor values* → *Spectra* → *Absorption* (→ p. 53).
Measurement signal must be approximately at zero and be straight as average value.

13 Fit pressure reducer on the test gas cylinder.



- ▶ Open the gas outlet of the pressure reducer and cylinder valve for a short time to flush the gas fitting.
- ▶ Set the pressure reducer to a value between 0.5 and 1 bar.
- ▶ Close the gas outlet and make a connection to the filter box.
- ▶ Open the gas outlet again afterwards.

14 Make a connection between fitting and filter box (test gas).

15 Set the valve to select the appropriate chamber or combination.

Set the “Measuring / Purging” valve to “Purging” to fill the chamber quickly.

16 When the measured value shown in SOPAS ET is constant, switch the “Measuring / Purging” to “Measuring”.

The overpressure is reduced.

Wait for a constant measured value and note the display.

17 Repeat test gas application with all possible chambers and combinations.

18 Close the cylinder valve. Switch valve “Measuring / Purging” to “Purging”.

Set all chambers to “Test gas”.

When the pressure in the pressure reducer has dropped to zero, switch all chamber valves to “Zero gas”.

19 Dismount the pressure reducer from the test gas cylinder.

To carry out another measurement with a different test gas:

Start again at → 13.

20 After measurement has completed, click “Maintenance”. The message “Maintenance (remove tool)” appears under “Operating mode”

21 Dismount the filter box from the GM32.

22 Fit the GM32 back onto the measuring point.

23 SOPAS ET menu *Maintenance* → *Operating mode* switch.

Click “Measuring” to switch the GM32 to operating mode “Measuring”.



The GM32 triggers a reference cycle.


The check is completed with operating mode “Measuring”.



The filter box scaling is overwritten automatically.

5.3 Test gas feeding (for GPP measuring probe)

Fig. 15 Test gas measurement with GPP probe




For GPP measuring probes, a gas check can be carried out with the measuring probe fitted in the gas duct.

An overpressure is created in the filter frit when gas (zero or test gas) is blown in. If this overpressure is high enough, the sample gas is forced out of the filter frit and replaced by the gas blown in.

The zero or test gas concentration in the filter frit is measured and the GM32 adjusted accordingly.

Fig. 16 Test gas connections on the GPP probe



- The GPP probe has a gas connection for test gas measurement.
- A second connection is available for, for example:
 - Pressure measurement
 - Sample extraction for comparison measurements
 - Filter backflushing
 - To increase the amount of test gas available.

5.3.1

One-off initial measurement - determining the test gas pressure

**WARNING: Hazard through escaping gases**

Hot and/or noxious gases can escape during work on the test gas connection, depending on the equipment conditions.

- Test gases may only be applied by skilled persons who, based on their technical training and knowledge as well as knowledge of the relevant regulations, can assess the tasks given and recognize the dangers involved.

Material required
SOPAS ET
Instrument air or N ₂
Pressure gauge / pressure reducer
Supply line 1/4" with Swagelok fitting

The GPP probe remains installed in the gas duct.

- 1 Switch the GM32 to *Maintenance* (on the operator panel or in SOPAS ET menu: *GM32/Maintenance/Operating mode switch* → p. 56, §4.2.6.1)
- 2 Connect N₂ or instrument air to the probe via a pressure reducer (→ p. 76, Fig. 16).
- 3 Set the pressure reducer to 0 bar, open the valve completely.
- 4 Set the pressure to approx. 3 bar.
- 5 Observe the measured values in the measured value display (operator panel or SOPAS ET).
The measured value display must show "Zero".
 When "Zero" is *not* displayed after approx. 30 s (depending on the T₉₀ setting of the analyzer): Increase the pressure gradually until "Zero" is displayed.
- 6 When "Zero" is displayed: Reduce the pressure in small steps until the measured value display starts to rise.
 Then increase the pressure by approx. 0.5 bar.
- 7 Note the pressure shown on the pressure gauge.
 Use this value as primary pressure value in the future.



Determine the test gas pressure again when the pressure in the gas duct changes.

The difference pressure is in the range 2 mbar ... 30 mbar.

- 8 Close the valve on the pressure reducer.
- 9 If no further measurements are to be made: Set the GM32 back into measuring operation (on the operating console or in SOPAS ET menu: *GM32/Maintenance/Operating mode switch*).

5.3.2

Manual test gas feeding**WARNING: Hazard through escaping gases**

Hot and/or noxious gases can escape during work on the test gas connection, depending on the equipment conditions.

- ▶ Test gases may only be applied by skilled persons who, based on their technical training and knowledge as well as knowledge of the relevant regulations, can assess the tasks given and recognize the dangers involved.

Material required
SOPAS ET
Instrument air or N ₂
Pressure gauge / pressure reducer
Supply line 1/4" with Swagelok fitting

- 1 Switch the GM32 to *Maintenance* (on the operator panel or in SOPAS ET menu: *GM32/Maintenance/Operating mode switch* → p. 56, §4.2.6.1)
- 2 Connect the desired zero and/or test gas via pressure reducer to the probe (→ p. 76, Fig. 16).
- 3 Set the pressure gauge (→ p. 77, §5.3.1) to the primary pressure determined.



Determine the test gas pressure again when the pressure in the gas duct changes.

- 4 If no further measurements are to be made: Set the GM32 back into measuring operation (on the operating console or in SOPAS ET menu: *GM32/Maintenance/Operating mode switch*).

5.3.3

Using mixed gases as test gases

Safe mixed conditions:

Mixed gas	Safe mixed gas conditions
SO ₂ + NO	> 10% NO included in SO ₂
SO ₂ + NO ₂	> 10% NO ₂ included in SO ₂
NO + SO ₂	> 10% SO ₂ included in NO
NO + NO ₂ ¹	> 25% NO ₂ included in NO
NO ₂ + SO ₂	> 50% SO ₂ included in NO ₂
NO ₂ + NO ¹	> 10% NO included in NO ₂

¹ As far as available

5.4

Fitting and removing the power supply unit of the connection unit

The removed power supply unit can be used for direct voltage supply for the SR-unit, e.g. when this must be removed from the gas duct for a zero adjust.

- 1 Disconnect the connection unit from the mains.
- 2 Open the housing cover of the connection unit (4 x 4 mm hexagon socket).

Fig. 17

Opening the connection unit



- 3 Loosen both plug connections of the power supply unit.
- 4 Unscrew the screw underneath the power supply unit (4 mm hexagon socket).
- 5 Push the power supply unit upwards and take out.
- 6 The power supply unit can be hung on the bracket of the SR-unit.

Fig. 18

Power supply unit hung on



Fit in the reverse sequence.

GM32

6 Shutting down/Restarting

6.1 Qualification requirements

The same qualifications as for maintenance are required for shutting down.

Correct handling of system components

During all work, ensure system components are not subjected to excessive mechanical or climatical stress:

- Always secure the SR-unit in a safe location and free from jolts.
- Protect system components against rain, dust or other unallowed ambient conditions, especially when dismantled.

6.2 Shutting down



Safety information for shutdown

- ▶ Wear suitable protective clothing and a protective mask when the sample gas is hot and/or aggressive or has a high dust load, or when the sample gas duct is under high pressure. Never open the housing, release the quick-release fasteners or remove the measuring probe without taking appropriate protective measures.
- ▶ If the conditions in the sample gas duct are particularly difficult and do not allow work on the open duct, even with protective equipment, the work must be carried out when the sample gas duct is inactive or has been flooded with ambient air.
- ▶ The purge air feed must remain in constant operation and the SR-unit may not be opened or swiveled as long as the measuring probe is in the sample gas duct. This is also applicable when the sample gas duct is out of operation because residual gases or dust can lead to contamination or corrosive deposits on the system components.
- ▶ As described in the following, to protect against IR radiation, first switch off the power supply to the SR-unit and disconnect the power supply plug before opening the housing or swiveling the flange fixture.

6.2.1 Equipment required

- Personal protective clothing and equipment (see safety information above)
- At least the following tools:
 - Complete wrench set containing, apart from others, the following wrenches:
 - 2 x 24 mm open-ended spanners or ring spanners
 - 1 x 19 mm open-ended spanner or ring spanner
 - Allen key set
 - Insulated screwdriver set
- Cleaning cloths for rough cleaning the outside of devices
- Consumables; when the measuring system is to be used again:
 - Desiccant cartridge (Part No.: 2 010 549) including the pin wrench provided to exchange the cartridges
- Suitable cover to close off the duct-side flange with tube after removing the measuring system (when necessary)
- Stored transport safety devices and, when available, the original packaging of the measuring system
- Insulation material/close box to protect cables against atmospheric conditions

6.2.2 Shutdown

- Ensure measuring signals are no longer evaluated.

6.2.2.1 Removing from the sample gas duct




Connections on GM32: See GM32 Operating Instructions

- If fitted, remove the weatherproof cover from the SR-unit by unscrewing the two knurled nuts on the top and lifting the cover from the SR-unit.
- Switch the power supply to the SR-unit off. Leave the purge air supply, if present, switched on for the time being.
- Unscrew the locking bushings of the three plug-in connectors on the underside of the SR-unit and remove the plug-in connectors.
- Loosen the cable clamp for the earthing conductor and remove the cable.
- If a GPP measuring probe is used, also remove as well the plug of the power supply cable from the heating control of the measuring probe.
- Under consideration of the safety information (→ p. 82, § 6.2), remove the SR-unit together with the fitted measuring probe out of the sample gas duct. To do this, loosen the four screw fastenings (→ p. 83, Fig. 19) with two 24 mm wrenches and pull the measuring probe with fitted SR-unit carefully out of the sample gas duct. On no account open the quick-release fasteners on the side of the SR-unit housing when the measuring probe with the SR-unit is still in the sample gas duct.

Fig. 19

Disconnecting the measuring probe with fitted SR-unit from the duct-side “flange with tube”



- Lay the SR-unit with the fitted measuring probe on the floor and free from jolts.

- ▶ Close off the duct-side flange with tube as necessary with a suitable protective cover.



High temperatures

If the sample gas duct is operated at high temperatures, the measuring probe will remain hot for some time after it has been removed.

- ▶ It should be handled with care in this case. GMP measuring probes cool down faster when the purge air supply remains switched on for a few minutes.

6.2.3

Disconnecting the purge air connection

On systems with purge air supply:

- ▶ Switch the power supply to the purge air unit off (as necessary, after an appropriate wait time for cooling down).
- ▶ Loosen the hose clamp of the purge air hose on the measuring probe or on the optional differential pressure sensor and take the purge air hose off.

If a differential pressure sensor is fitted (see Fig. 7.10, page 80):

- ▶ Loosen the hose clamp on the purge air inlet of the measuring probe and take off the differential pressure sensor including the associated hose piece.
- ▶ Disconnect the signal cable on the differential pressure sensor by pulling off both 6.3 mm flat pin bushings in the housing.
- ▶ If the signal cable of the existing measuring point should be reused, protect the cable end with the open flat pin bushings against atmospheric conditions with insulation material or a close box.

The next tasks are to first put the purge air unit and further components - as far as applicable - out of operation.

6.2.4

Putting the purge air unit out of operation

On systems with purge air supply:



WARNING: Ensure the power supply remains switched off

- ▶ Ensure the power supply of the purge air unit is completely switched off and remains so by switching off the corresponding circuit breaker and the motor switch, and secure these against being switched on again.

If the existing measuring point is to be used for the purge air unit again, e.g., for a different measuring system:

- ▶ Check the technical data or contact SICK to determine whether the purge air unit is suitable for the planned application purpose.
 - ▶ If yes: Close off the open end of the purge air hose tight to prevent dirt or moisture entering the hose.

If no further use on the measuring point is planned:

- ▶ Remove the weatherproof cover, if fitted, from the purge air unit.
- ▶ Open the connection unit, loosen the connection cable clamps and remove the cable.
- ▶ If the connection cable is to remain at the measuring point, insulate the open cable end and protect them with a close box against atmospheric conditions.
- ▶ Loosen the fastening screws of the purge air unit and take it off from the retainers or base as completely assembled unit
- ▶ If desired, remove the purge air hose. To do this, loosen its hose clamp on the outlet connection of the purge air unit.

6.3 Further components on the measuring point

6.3.1 Connection unit and cabling

If the existing cabling at the measuring point is no longer required in its present form, remove the cabling as described in the following.

- ▶ Disconnect all lines provided by the customer from the terminal strip of the connection unit.
- ▶ Loosen the fastening of the connection unit.
- ▶ Remove the connection unit with all pre-assembled cables with plugs.
- If there are problems in removing the pre-assembled cables with plugs together with the connection unit:
 - ▶ Disconnect the pre-assembled cables from the terminal strip in the connection unit and remove these from the installation ducts or installation location.

If the cables provided by the customer at the measuring point can be used for other purposes:

- ▶ Insulate all open cable ends.
- ▶ Protect the cable ends against atmospheric conditions, e.g., with a close box to prevent corrosion and contamination.

6.4

Conservation measures, correct storage and transport

This Section describes how to prepare the GM32 for storage for later use or transport.

- ▶ Clean the outside of the closed SR-unit housing and all other components with damp cleaning cloths. Pay special attention to removing chemical residues or condensate deposits. A mild cleaning agent that does not leave any residues can be used here.

- ▶ Select a dry, frost-free room without strong temperature fluctuations for storage.

If the original packaging is available or in all cases where the SR-unit and measuring probe are to be stored or transported separately:

- ▶ Separate the measuring probe from the SR-unit (→ p. 83, Fig. 19):
 - ▶ Push the sealing ring to one side so that the measuring probe can be removed without hindrance after the screw fitting has been loosened.
 - ▶ Loosen the three nuts with a 19 mm wrench and carefully remove the measuring probe without damaging the threaded bolts on the SR-unit.
- ▶ Keep all the fastening parts with care.

To prevent absorbed moisture damaging the SR-unit, check and replace the desiccant cartridges when necessary:

- ▶ Check the desiccant cartridges. As preparation for the storage period, the cartridges should then also be exchanged when the drying granulate is only slightly discolored when compared with new cartridges.
- ▶ Protect the openings on the SR-unit and measuring probe against atmospheric conditions. In this case, use the transport safety devices, as far as available, used during delivery of these components. Apart from that, materials such as weatherproof adhesive tape and plastic foils are also suitable.
- ▶ If the measuring probe and SR-unit remain connected, the front cover of the SR-unit and the plastic protective cap on the measuring probe flange are not required. If a GPP measuring probe is used and is to remain fitted, no transport safety devices are required because GPP measuring probes do not have an open measuring gap.


Fig. 20

Transport safety devices in the SR-unit



Fig. 21

Transport safety devices on the probe (shown here on a GPP probe)



- Pack the complete measuring system for storage or transport preferably in the original packaging. If this is not available, ensure the packaging used, under consideration of the weight of system components as well, provides adequate protection against impact loads as well as against atmospheric conditions.

6.5 **Preparing for restarting**

6.5.1 **New measuring task**

If the planned measuring task is different to the original task (e.g., modified sample gas composition, different measuring point):

- ▶ Ask SICK Service or the responsible sales partner whether a new factory calibration is required for the changed measuring task and which system components may have to be returned (usually the SR-unit).

If a factory calibration is necessary:

- ▶ If the measuring system has not been stored in the meantime, prepare it for transport as described in §6.4 (→ p. 87).
- ▶ The original packaging is best suited for shipping. The packaging must always provide sufficient protection for transport.
- ▶ Send the SR-unit and/or other system components as agreed to the address specified by your SICK contact person.

6.5.2 **Previous measuring task**

If a factory calibration is not necessary for the new measuring task and the measuring system is used for the new measuring task without intermediate storage:

- ▶ Clean the outside of the closed SR-unit housing and the connected measuring probe with damp cleaning cloths. Pay special attention to removing chemical residues or condensate deposits. A mild cleaning agent that does not leave any residues can be used here.

6.5.3 **Purge air unit maintenance**

On measuring systems with purge air unit:

- ▶ Perform maintenance on the purge air unit before restarting.

GM32

7 Annex

Commissioning checklist

7.1

Commissioning checklist

Commissioning Check List GM32			
Customer data			
Customer: _____	Customer no: _____		
Country: _____	City: _____		
Plant: _____	Location: _____		
1 Device data			
Device type: _____	Device no: _____		
Serial no: _____			
Process optic: GMP ☐	GPP ☐	CD ☐	
Serial no: _____	Device no: _____	Type: _____	
2 Plant data			
Location: _____	Outside ☐	Under cover ☐	Inside ☐
Tag number _____			
Orientation of the stack _____	Horizontal ☐	Vertical ☐	Angle of _____ °
Orientation of the GM32 _____	Horizontal ☐	Vertical ☐	Angle of _____ °
Length of probe / Flange-Flange (CD) _____ mm	Active measurement distance / Aperture _____ mm		
Zero path _____ mm	Differential pressure _____ hpa		
Ambient temperature _____ °C	Gas temperature _____ °C		
I/O Modules on site ☐	I/O Modules relocated ☐		
Plant operating status _____			
3 Prerequisite			
	Y	N	Remarks
3.1 Documentation + Delivery complete	☐	☐	
3.2 Platform at measurement spot has suitable dimension?	☐	☐	
3.3 If this measurement location is under legal regulation, has it been acknowledged by an official body?	☐	☐	
3.4 Customer specific data for parameterization available?	☐	☐	
3.5 Purge air unit installed and electrically connected?	☐	☐	
3.6 Connection unit installed and electrically connected?	☐	☐	
3.7 Zero point stands / tube available (only CD)?	☐	☐	

4 Preliminary work		Y	N	Remarks
4.1	Mounting of flanges like described in the Operating Instruction?	☐	☐	
4.2	Check for damage	☐	☐	
4.3	Check ambient conditions (ref. ch. 2)	☐	☐	
4.4	Check mounting conditions (ref. ch. 2)	☐	☐	
4.5	Check mounting	☐	☐	
4.6	Check cables / wires for correct installation	☐	☐	
4.7	Check main power supply voltage	☐	☐	

5 Purge Air Unit (only GMP / Cross-Duct)		Y	N	Remarks
5.1	Purge air unit type	X	X	
5.2	Check the rotation direction	☐	☐	
5.3	Check hoses for correct installation	☐	☐	
5.4	Purge air heating installed?	☐	☐	
5.5	Differential pressure monitor installed?	☐	☐	

6 Sender / receiver unit		Y	N	Remarks
6.1	Clean all optical surfaces	☐	☐	
6.2	Dessicant cartridges exchanged?	☐	☐	
6.3	Check output of power supply voltages	☐	☐	
6.3.1	26V±0,3V:	☐	☐	
6.3.2	115V±1V:	☐	☐	
6.4	Check emission free zero point	☐	☐	
6.5	Adjust purge air heating (option)	☐	☐	Value:
6.6	Check and adjust differential pressure monitor (option)	☐	☐	
6.7	Install the GM32 at the measurement spot	☐	☐	
6.8	Adjust optical alignment	☐	☐	
6.9	Check signals	☐	☐	
A:		B:		
C:		D:		
6.10	Note lamp data	☐	☐	
Maximum lamp intensity:		Exposure:		ms
		Lamp pulse:		mA
6.11	Note software revision	☐	☐	
Device Process:		Display:		
Process optic:		/		

6.12	Check stepper motor for proper function	☐	☐
6.13	Check parameterization	☐	☐
6.14	Parameterize the I/O Modules	☐	☐
6.15	Perform loop test	☐	☐
6.16	Measured values are plausible (ref. Ch. 8)	☐	☐
6.17	Save device data	☐	☐
6.18	Complete Commissioning Sign-Off Sheet	☐	☐
6.19	Instruct the operator personnel Hand over the maintenance manual and check lists - Measurement reading - Perform customer maintenance - Read messages	☐	☐

7 Input / Output / Ranges

7.1 Analog output / Display

Live zero: mA			Output of control values: Y ☐ / N ☐				
Analog output	Comp.	Source	Unit	Range		Control value	Remark
				Start	Ende		
1						☐	
2						☐	
3						☐	
4						☐	
5						☐	
6						☐	
7						☐	
8						☐	
9						☐	
10						☐	
11						☐	
12						☐	
13						☐	
14						☐	
15						☐	
16						☐	

7.2 Analog input

Analog input	Source	Unit	Live zero	Start	End
1					
2					

7.3. Digitalausgang

Digitalausgang	Quelle	Inv.
1		☐
2		☐
3		☐
4		☐

7.4 Digital input

Digital input	Source	Inv.
1		☐
2		☐
3		☐
4		☐

8 Measurement values after commissioning

	Unit	Range	Reading	Zero point	Span point
SO ₂					
NO					
Temp.				- X -	- X -
Press.				- X -	- X -

Remarks

--

Date : _____	Name Plant personnel: _____ Engineer: _____
--------------	---

A		
Absorption		
- Spectrum	53	
- Filtered	53	
- Wavelength corrected	54	
Adjustment (menu)	42	
Adjustment sensor (menu)	36, 48	
Alignment (menu)	42	
Analog inputs (parameter setting)	38	
Analog outputs (parameter setting)	38	
Aperture signal (dark signal) (spectrum)	52	
Authorized operator	32	
B		
Bargraphs	33	
Boxmeasuring		
- Menu	44	
- Procedure	68	
C		
C (classification)	46	
CAN		
- CAN bus connection	24	
- Connection	22	
Cell filling (monitoring)	14	
Check cycle	47	
- Description	12	
- Maintenance mode	56	
- Signals	15	
Checklist		
- Commissioning	92	
Classification	46	
Coefficients	35 - 36	
Coefficients boxmeasuring	36	
Commissioning		
- Checklist	92	
Contamination compensation	16	
D		
Data files, loading from device	63	
Device description file, loading	61	
Device information (menu)	45	
Device parameter (menu)	34	
Digital inputs (configuration)	39	
Digital outputs (configuration)	39	
E		
E (classification)	46	
EEPROM		
- Loading	57	
- Save	57	
F		
F (classification)	46	
Functional principle	10	
H		
Hardware Map	41	
I		
I/O module	41	
IP address	59	
L		
Logbook	45	
- Defaults	37	
- Export	46	
M		
M (classification)	46	
Maintenance	65	
Maintenance (switching operating mode)	56	
Maintenance mode	56	
Maintenance request	46	
Maintenance request (signal)	15	
Mapping Table (SCU)	26	
Measured value (menu)	33	
Measurement reflector (spectrum)	51	
Measuring (switching operating mode)	56	
Measuring range switch-over	38	
Menu tree	31	
Mirror tracking	16	
- Menu	49	
N		
Normalisation (spectrum)	55	
O		
Operation	29	
Optics layout	10	
P		
Parameters		
- Loading	62	
- Menu	34	
- Saving	62 - 63	
Password	32	
Pressure (analog input)	38	
Process optic (menu)	49 - 50	
Product description	9	
Program update	60	
Q		
QAL3	12	

R

Reference cycle	
- Description	12
- Maintenance mode	56
Reference point	
- Determining	14
- Output	15
Reference spectrum	54
Reset	56
Resolution drift (monitoring)	13

S

SCU	30
- Connection on SCU	24
- Mapping Table	26
SDD file	58
sdv file	58
Sensor values	47
Service level	32
Signal evaluation	11
smf file	58
SOPAS ET	
- Description	30
- First steps	30
- Menu tree	31
Spectra	51
Spectral analysis (menu)	35
Spectrometer (menu)	35, 48
Start-up Assistant	57
System reset	56

T

Target group	8
Temperature (analog input)	38
Terminating resistor	22
- Parameter setting	34

U

U (classification)	46
Updating firmware	60
User	8
User level	32

W

Wavelength scale (monitoring)	13
-------------------------------------	----

Z

Zero adjust	
- Description	12
- Menu	43
- Procedure	66
Zero point	
- Determining	14
- Output	15
Zero point reflector (spectrum)	52

8030993/AE00/V1-4/2013-11

www.addresses.endress.com

For more Endress+Hauser product please visit: www.ainstru.com
Mail: Sales@ainstru.com