

Operating Instructions

Conductive multiple rod level switch for liquids

VEGAKON 66

Relay (DPDT)



Document ID: 32649



VEGA

Contents

1	About this document	4
1.1	Function	4
1.2	Target group	4
1.3	Symbols used.....	4
2	For your safety	5
2.1	Authorised personnel	5
2.2	Appropriate use.....	5
2.3	Warning about incorrect use.....	5
2.4	General safety instructions	5
2.5	Safety label on the instrument	6
2.6	EU conformity.....	6
2.7	Installation and operation in the USA and Canada	6
2.8	Environmental instructions	6
3	Product description	7
3.1	Configuration.....	7
3.2	Principle of operation.....	8
3.3	Adjustment	9
3.4	Storage and transport.....	9
4	Mounting.....	11
4.1	General instructions	11
4.2	Mounting instructions	12
5	Connecting to power supply.....	13
5.1	Preparing the connection	13
5.2	Connection instructions.....	13
5.3	Wiring plan	14
6	Setup	15
6.1	General information.....	15
6.2	Adjustment elements.....	15
6.3	Switching point adjustment	16
6.4	Function table.....	19
7	Maintenance and fault rectification.....	20
7.1	Maintenance	20
7.2	Electronics exchange	20
7.3	Simulation of switching functions	21
7.4	Shortening the probe.....	21
7.5	How to proceed if a repair is necessary	21
8	Dismount.....	23
8.1	Dismounting steps.....	23
8.2	Disposal	23
9	Supplement	24
9.1	Technical data	24
9.2	Dimensions	26
9.3	Industrial property rights.....	27
9.4	Trademark	27

1 About this document

1.1 Function

This operating instructions provides all the information you need for mounting, connection and setup as well as important instructions for maintenance, fault rectification, the exchange of parts and the safety of the user. Please read this information before putting the instrument into operation and keep this manual accessible in the immediate vicinity of the device.

1.2 Target group

This operating instructions manual is directed to trained personnel. The contents of this manual must be made available to the qualified personnel and implemented.

1.3 Symbols used



Document ID

This symbol on the front page of this instruction refers to the Document ID. By entering the Document ID on www.vega.com you will reach the document download.



Information, tip, note

This symbol indicates helpful additional information.



Caution: If this warning is ignored, faults or malfunctions can result.



Warning: If this warning is ignored, injury to persons and/or serious damage to the instrument can result.



Danger: If this warning is ignored, serious injury to persons and/or destruction of the instrument can result.



Ex applications

This symbol indicates special instructions for Ex applications.



SIL applications

This symbol indicates instructions for functional safety which must be taken into account particularly for safety-relevant applications.

- **List**

The dot set in front indicates a list with no implied sequence.

- **Action**

This arrow indicates a single action.

- 1 **Sequence of actions**

Numbers set in front indicate successive steps in a procedure.



Disposal

This symbol indicates special instructions for disposal.

2 For your safety

2.1 Authorised personnel

All operations described in this documentation must be carried out only by trained, qualified personnel authorised by the plant operator.

During work on and with the device, the required personal protective equipment must always be worn.

2.2 Appropriate use

The VEGAKON 66 is a sensor for point level detection.

You can find detailed information about the area of application in chapter " *Product description*".

Operational reliability is ensured only if the instrument is properly used according to the specifications in the operating instructions manual as well as possible supplementary instructions.

For safety and warranty reasons, any invasive work on the device beyond that described in the operating instructions manual may be carried out only by personnel authorised by the manufacturer. Arbitrary conversions or modifications are explicitly forbidden.

2.3 Warning about incorrect use

Inappropriate or incorrect use of this product can give rise to application-specific hazards, e.g. vessel overfill through incorrect mounting or adjustment. Damage to property and persons or environmental contamination can result. Also, the protective characteristics of the instrument can be impaired.

2.4 General safety instructions

This is a state-of-the-art instrument complying with all prevailing regulations and directives. The instrument must only be operated in a technically flawless and reliable condition. The operator is responsible for the trouble-free operation of the instrument. When measuring aggressive or corrosive media that can cause a dangerous situation if the instrument malfunctions, the operator has to implement suitable measures to make sure the instrument is functioning properly.

The safety instructions in this operating instructions manual, the national installation standards as well as the valid safety regulations and accident prevention rules must be observed by the user.

For safety and warranty reasons, any invasive work on the device beyond that described in the operating instructions manual may be carried out only by personnel authorised by the manufacturer. Arbitrary conversions or modifications are explicitly forbidden. For safety reasons, only the accessory specified by the manufacturer must be used.

To avoid any danger, the safety approval markings and safety tips on the device must also be observed.

2.5 Safety label on the instrument

The safety approval markings and safety tips on the device must be observed.

2.6 EU conformity

The device fulfils the legal requirements of the applicable EU directives. By affixing the CE marking, we confirm the conformity of the instrument with these directives.

The EU conformity declaration can be found on our homepage.

2.7 Installation and operation in the USA and Canada

This information is only valid for USA and Canada. Hence the following text is only available in the English language.

Installations in the US shall comply with the relevant requirements of the National Electrical Code (ANSI/NFPA 70).

Installations in Canada shall comply with the relevant requirements of the Canadian Electrical Code.

2.8 Environmental instructions

Protection of the environment is one of our most important duties. That is why we have introduced an environment management system with the goal of continuously improving company environmental protection. The environment management system is certified according to DIN EN ISO 14001.

Please help us fulfil this obligation by observing the environmental instructions in this manual:

- Chapter " *Packaging, transport and storage* "
- Chapter " *Disposal* "

3 Product description

3.1 Configuration

Scope of delivery

The scope of delivery encompasses:

- VEGAKON 66 compact level switch
- Documentation
 - This operating instructions manual

Constituent parts

The VEGAKON 66 consists of the components:

- Housing lid
- Housing with electronics
- Process fitting with electrodes




Fig. 1: VEGAKON 66

- 1 Housing lid
- 2 Housing with electronics
- 3 Process fitting with electrodes

Serial number - Instrument search

The type label contains the serial number of the instrument. With it you can find the following instrument data on our homepage:

- Product code (HTML)
- Delivery date (HTML)
- Order-specific instrument features (HTML)
- Operating instructions and quick setup guide at the time of shipment (PDF)
- Order-specific sensor data

Move to "www.vega.com" and enter in the search field the serial number of your instrument.

Alternatively, you can access the data via your smartphone:

- Download the VEGA Tools app from the "*Apple App Store*" or the "*Google Play Store*"
- Scan the DataMatrix code on the type label of the instrument or
- Enter the serial number manually in the app

3.2 Principle of operation

Application area

The conductive VEGAKON 66 compact level switches detect levels of conductive liquids.

Functional principle

If at least two electrodes are covered with a conductive medium, small alternating currents ($< 1 \text{ mA}$) flow from the measuring electrode to the ground electrode.




Fig. 2: Functional principle

- 1 Ground probe
- 2 Max. level (max. electrode)
- 3 Min. level (min. electrode)

These alternating currents are measured in respect to their amplitude and phase position and converted into a switching command.

VEGAKON 66 can be used for reliable detection of products over a very wide conductivity and viscosity range.

A level detection can be realised with two measuring electrodes, e.g. a pump or two-point control with three measuring electrodes.

Voltage supply

VEGAKON 66 is a compact instrument, i.e. it can be operated without external evaluation system. The integrated electronics evaluates the level signal and outputs a switching signal. With this switching signal, a connected device can be operated directly (e.g. a warning system, a pump etc.).

The data for power supply are specified in chapter " *Technical data*".

3.3 Adjustment

The VEGAKON 66 is a compact level switch with integrated electronics module.

On the electronics module you will find the following display and adjustment elements:

- Control lamp for indication of the switching status
- Mode changeover for selection of the output signal
- DIL switch for adjustment of the damping
- Rotary switch for adjustment of the conductivity value

3.4 Storage and transport

Packaging

Your instrument was protected by packaging during transport. Its capacity to handle normal loads during transport is assured by a test based on ISO 4180.

The packaging of standard instruments consists of environment-friendly, recyclable carton material. The sensing element is additionally protected with a cardboard cover. For special versions, PE foam or PE foil is also used. Please dispose of the packaging material through specialised recycling companies.

Transport

Transport must be carried out in due consideration of the notes on the transport packaging. Nonobservance of these instructions can cause damage to the device.

Transport inspection

The delivery must be checked for completeness and possible transit damage immediately at receipt. Ascertained transit damage or concealed defects must be appropriately dealt with.

Storage

Up to the time of installation, the packages must be left closed and stored according to the orientation and storage markings on the outside.

Unless otherwise indicated, the packages must be stored only under the following conditions:

- Not in the open
- Dry and dust free
- Not exposed to corrosive media
- Protected against solar radiation
- Avoiding mechanical shock and vibration
- Storage and transport temperature see chapter " *Supplement - Technical data - Ambient conditions*"
- Relative humidity 20 ... 85 %

Storage and transport temperature

Lifting and carrying

With instrument weights of more than 18 kg (39.68 lbs) suitable and approved equipment must be used for lifting and carrying.

4 Mounting

4.1 General instructions

Process conditions



Note:

For safety reasons, the instrument must only be operated within the permissible process conditions. You can find detailed information on the process conditions in chapter " *Technical data*" of the operating instructions or on the type label.

Hence make sure before mounting that all parts of the instrument exposed to the process are suitable for the existing process conditions.

These are mainly:

- Active measuring component
- Process fitting
- Process seal

Process conditions in particular are:

- Process pressure
- Process temperature
- Chemical properties of the medium
- Abrasion and mechanical influences

Suitability for the ambient conditions

The instrument is suitable for standard and extended ambient conditions acc. to DIN/EN/IEC/ANSI/ISA/UL/CSA 61010-1. It can be used indoors as well as outdoors.

Moisture

Use the recommended cables (see chapter " *Connecting to power supply*") and tighten the cable gland.

You can give your VEGAKON 66 additional protection against moisture penetration by leading the connection cable downward in front of the cable gland. Rain and condensation water can thus drain off. This applies mainly to outdoor mounting as well as installation in areas where high humidity is expected (e.g. through cleaning processes) or on cooled or heated vessels.

To maintain the housing protection, make sure that the housing lid is closed during operation and locked, if necessary.




Fig. 3: Measures against moisture ingress

Pressure/Vacuum

The process fitting must be sealed if there is gauge or low pressure in the vessel. Before use, check if the sealing material is resistant against the measured product and the process temperature.

The max. permissible pressure is specified in chapter " *Technical data*" or on the type label of the sensor.

**Cable entries - NPT
thread
Cable glands**
Metric threads

In the case of instrument housings with metric thread, the cable glands are screwed in at the factory. They are sealed with plastic plugs as transport protection.

You have to remove these plugs before electrical connection.

NPT thread

In the case of instrument housings with self-sealing NPT threads, it is not possible to have the cable entries screwed in at the factory. The free openings for the cable glands are therefore covered with red dust protection caps as transport protection.

Prior to setup you have to replace these protective caps with approved cable glands or close the openings with suitable blind plugs.

Agitators**4.2 Mounting instructions**

Due to agitators, equipment vibration or similar, the level switch can be subjected to strong lateral forces.

During operation, the probe must not touch any installations or the vessel wall. If necessary, secure the end of the probe (insulated). Provide a suitable insulated support directly above the probe end.




Fig. 4: Fasten the probe

- 1 Measuring probe
- 2 Plastic sleeve at the probe end or laterally mounted

5 Connecting to power supply

5.1 Preparing the connection

Note safety instructions

Always keep in mind the following safety instructions:



Warning:

Connect only in the complete absence of line voltage.

- The electrical connection must only be carried out by trained, qualified personnel authorised by the plant operator.
- Always switch off power supply, before connecting or disconnecting the instrument.



Note:

Install a disconnecting device for the instrument which is easy to access. The disconnecting device must be marked for the instrument (IEC/EN 61010).

Voltage supply

Connect the voltage supply according to the following connection diagrams. The electronics module KONE60R is designed in protection class I. To maintain this protection class, it is absolutely necessary that the ground conductor be connected to the internal ground terminal. Take note of the general installation regulations.

The data for power supply are specified in chapter " *Technical data*".

Connection cable

The instrument is connected with standard three-wire cable without shielding. If electromagnetic interference is expected which is above the test values of EN 61326 for industrial areas, shielded cable should be used.

Make sure that the cable used has the required temperature resistance and fire safety for max. occurring ambient temperature

Use cable with round cross-section. A cable outer diameter of 5 ... 9 mm (0.2 ... 0.35 in) ensures the seal effect of the cable gland. If you are using cable with a different diameter or cross-section, exchange the seal or use a suitable cable gland.

5.2 Connection instructions



Danger:

Switch off power supply before starting connection work.

Connect mains voltage according to the connection diagrams.

Marking of the probe rods

The probe rods are marked with numbers. You can find the numbers on the process fitting below the thread.




Fig. 5: Numbering of the probes - Bottom view

- 1 Ground rod - has the same length or is longer than the min. rod
- 2 Max. rod - its length defines the upper switching level (shortest rod)
- 3 Min. rod - its length defines the lower switching level (mid rod)

5.3 Wiring plan

Floating relay output

Is used to switch external voltage sources to relays, contactors, magnetic valves, horns etc.




Fig. 6: Electronics with relay output

- 1 Relay output
- 2 Voltage supply

Connection to a PLC

If inductive loads or stronger currents are switched through, the gold plating on the relay contact surface will be permanently damaged. The contact is then no longer suitable for switching low-voltage circuits.

Inductive loads also result from the connection to a PLC input or output and/or in combination with long cables. It is imperative that you take measures to extinguish sparks to protect the relay contact (e.g. Z diode) or use an electronic version with transistor output.

6 Setup


6.1 General information

Function/Configuration

On the electronics module you will find the following display and adjustment elements:

- DIL switch for mode adjustment
- DIL switch for adjustment of the damping
- Rotary switch for adjustment of the conductivity value
- Control lamp for indication of the switching status

6.2 Adjustment elements



- 1 Connection terminals
- 2 Control lamp (LED)
- 3 Rotary switch: Adjustment of the conductivity value
- 4 Selection switch: Switch-on delay
- 5 Selection switch: Mode (A/B) VEGAKON 66
- 6 Type label
- 7 Tensile proving ring

Signal lamp (2)

The switching condition of the signal lamp can be checked when the housing is closed. To adjust VEGAKON 66 loosen the four screws with a wrench on the upper side of the instrument and remove the housing cover.

Rotary switch: Adjustment of the conductivity value (3)

With the rotary switch you can set the sensitivity of the instrument. Position 0.1 kΩ is the less sensitive and switch position 300 kΩ the most sensitive one, see table "Switching point adjustment".

Selection switch: Switching delay (4)

There are three switches on the DIL switch block by which you can adjust the switching on and off delay. Thus prevents, e.g. the instrument from permanent switching if the level is within the limits.

The switching delay refers to the switching status of both relay outputs.

With the switches (2 s, 6 s, 12 s) you can set the switching delay respectively in the range of 0 to 20 seconds. The times of the activated timers add up. If e.g. the switches 2 s and 12 s are activated, the switching delay will be 14 s.

Mode adjustment (5)

With the mode adjustment (A/B) you can change the switching condition of the output. You can set the required mode according to the "Function table" (A - max. detection or overflow protection, B - min. detection or dry run protection).

Tensile proving ring (7)

Loosen the holding screws of the electronics module. Fold the tensile proving ring upward. With the tensile proving ring you can pull the electronics module out of the instrument housing.

6.3 Switching point adjustment**Rotary switch position**

Switching point at approx. 1 cm covering.

Rotary switch position	Conductivity value (medium)
Test max.	Switching condition with complete covering is simulated
0.1 k Ω (insensitive)	> 6.6 mS
0.3 k Ω	> 1.7 mS
1 k Ω	> 540 μ S
3 k Ω	> 180 μ S
10 k Ω	> 54 μ S
30 k Ω	> 20 μ S
100 k Ω	> 5.7 μ S
300 k Ω (sensitive)	> 1.6 μ S
Test min.	Empty state is simulated

Examples of conductivity values

Medium	Conductivity value	Recommended rotary switch position
Tap water	0.2 mS	3 k Ω
Saltwater (3.5 %)	35 mS	0.1 k Ω
Beer	1.4 mS	1 k Ω
Fruit juice	2 mS	0.3 k Ω
Milk, yogurt	3 mS	0.3 k Ω
Ketchup	15 mS	0.1 k Ω

Determination of the response height

In case of horizontally mounted instruments, the response height is determined by the mounting height.

In case of vertically mounted instruments, the response height is determined by the length of the probe rods.

The modification of the response height by turning the conductivity value switch is not useful.

To adapt the response height, you can shorten the measuring rods by sawing them off. Before shortening, unscrew the rods out of the plastic threaded part of the sensor to avoid damaging the probe.

The numbers of the rods are visible on the lower side of the thread.

The ground rod (no. 1) must have the same length or must even be longer than the longest of the other rods.

The max. rod (no. 2) defines the response height with single point level switches or the upper switching level with two-point control. It is the shortest rod.

The min. rod (no. 3) defines the lower switching level, it must hence be longer than the max. rod. It is not available with instruments for single point level detection.

The electronics recognises if a min. rod is screwed in and changes automatically from single to two-point control.

Standard setting

For products with a good conductivity ($> 3 \text{ mS}$), set the rotary switch - conductivity setting (3) generally to switch position $3 \text{ k}\Omega$. Hence the instrument is already completely adjusted.

Take note of the instructions in table "Examples of conductivity values". The recommended settings take also influences such as e.g. condensation or slight buildup into account.

If there is danger of strong buildup or condensation, you can set the instrument more insensitive by one rotary switch position.

The "Function table" shows how to select the switching condition of the output.

Level detection for max. signal

The following settings apply to products with low conductivity ($< 3 \text{ mS}$)

1. Fill the vessel until the shortest probe is covered approx. 1 cm
2. Switch on the power supply
3. Set the A/B switch to mode A
4. Set the rotary switch to position "TEST min."
5. Turn the rotary switch slowly clockwise until the red LED lights

The instrument is adapted to the medium, i.e. the relay deenergized with max. level.

Level detection for min. signal

The following settings apply to products with low conductivity ($< 3 \text{ mS}$)

1. Empty the vessel until the min. measuring electrode is only covered approx. 1 cm
2. Switch on the power supply

3. Set the A/B switch to mode B
 4. Set the rotary switch to position "TEST max."
 5. Turn the rotary switch slowly anticlockwise until the red LED lights
- The instrument is adapted to the medium, i.e. the relay deenergized with min. level.

Two-point control - A-mode

The following settings apply to products with low conductivity (< 3 mS)

1. Fill the vessel until the shortest probe is covered approx. 1 cm
2. Switch on the power supply
3. Set the A/B switch to mode A
4. Set the rotary switch to position "TEST min."
5. Turn the rotary switch slowly clockwise until the red LED lights

The instrument is adapted to the medium, i.e. the relay deenergizes when the max. level is reached. Only when the min. probe is decreased, the relay energizes again.

Example: A filling pump is switched on when the level falls below the min. signal, fills the vessel until the max. signal is reached and is then switched off.

Two-point control - B-mode

The following settings apply to products with low conductivity (< 3 mS)

1. Empty the vessel until the min. measuring electrode is only covered approx. 1 cm
2. Switch on the power supply
3. Set the A/B switch to mode B
4. Set the rotary switch to position "TEST max."
5. Turn the rotary switch slowly anticlockwise until the red LED lights

The instrument is adapted to the medium, i.e. the relay energizes at max. level. Only of the min. probe is decreased, the relay deenergizes again.

Example: An emptying pump is switched on when the level reaches the max. signal, empties the vessel until the min. signal is reached and is then switched off.

Dry adjustment

In case of several measurement loops (same medium), the adjustment of one instrument with medium is sufficient. The determined switch position can be transferred to all other instruments.

When the conductivity value is known, the switching point adjustment can be carried out according to the setup in table " *Rotary switch adjustment*".

When exchanging the electronics module, it is sufficient to take over the setting of the old electronics module.

6.4 Function table

The following table provides an overview of the switching conditions depending on the set mode and the level.

	Level	Switching status, relay module E60R	Control lamp
Mode A Overflow protection	 ①	Relay energized 	 does not light
Mode A Overflow protection	 ②	Relay deenergized 	 lights
Mode B Dry run protection	 ③	Relay energized 	 does not light
Mode B Dry run protection	 ④	Relay deenergized 	 lights
Failure of the supply voltage (mode A/B)		Relay deenergized 	 does not light

1 Max. detection - Vessel empty

2 Max. detection - Vessel full

3 Min. detection - Vessel full

4 Min. detection - Vessel empty



Note:

If VEGAKON 66 is used for oil warning in water, the probe must be cleaned after having responded to oil (= empty signal) because otherwise resetting to water will not be ensured reliably.

7 Maintenance and fault rectification

7.1 Maintenance

Maintenance

If the device is used properly, no special maintenance is required in normal operation.

Cleaning

The cleaning helps that the type label and markings on the instrument are visible.

Take note of the following:

- Use only cleaning agents which do not corrode the housings, type label and seals
- Use only cleaning methods corresponding to the housing protection rating

7.2 Electronics exchange

In general, all electronics modules of series KONE66 can be interchanged. If you want to use an electronics module with a different signal output, you can download the corresponding operating instructions manual from our homepage under Downloads.

Proceed as follows:

1. Switch off voltage supply
2. Unscrew the housing lid
3. Loosen compression fittings with a screwdriver
4. Pull the connection cables out of the terminals
5. Loosen the two screws with a screw driver (Phillips recessed head)
6. Lift tensile proving ring and pull out the old electronics module
7. Compare the new electronics module with the old one. The type label of the electronics module must correspond to that of the old electronics module.
8. Note settings of all adjustment elements of the old electronics module.

Set the adjustment elements of the new electronics module to the same settings of the old one.

9. Screw in and tighten the two holding screws with a screwdriver (Phillips)
10. Insert the wire ends into the open terminals according to the wiring plan
11. Tighten the screw terminals
12. Check the hold of the wires in the terminals by lightly pulling on them
13. Check cable gland on tightness. The seal ring must completely encircle the cable.
14. Screw the housing lid back on

The electronics exchange is now finished.

As soon as you insert the electronics module, VEGAKON 66 is ready for operation.

7.3 Simulation of switching functions

With the rotary switch for conductivity adjustment, full covering or empty signal can be simulated.

The level must not be changed. You can hence easily check the response of connected signalling and switching facilities. Parts of the sensor electronics are also checked during this test.

The following switch positions, simulate the switching conditions:

- Switch position " *Test max.*" full covering (max.)
- Switch position " *Test min.*" empty signal (min.)

7.4 Shortening the probe

The rods of the probe can be shortened.

Marking of the probe rods The probe rods are marked with numbers. You can find the numbers on the process fitting below the thread.




Fig. 7: Numbering of the probes - Bottom view

- 1 Ground rod - has the same length or is longer than the min. rod
- 2 Max. rod - its length defines the upper switching level (shortest rod)
- 3 Min. rod - its length defines the lower switching level (mid rod)



Note:

Take note when shortening meas. rods that terminal 1 corresponds to the longest rod and terminal 2 to the shortest rod.

7.5 How to proceed if a repair is necessary

You can find an instrument return form as well as detailed information about the procedure in the download area of our homepage. By doing this you help us carry out the repair quickly and without having to call back for needed information.

In case of repair, proceed as follows:

- Print and fill out one form per instrument
- Clean the instrument and pack it damage-proof
- Attach the completed form and, if need be, also a safety data sheet outside on the packaging

- Ask the agency serving you to get the address for the return shipment. You can find the agency on our homepage.

8 Dismount

8.1 Dismounting steps

**Warning:**

Before dismantling, be aware of dangerous process conditions such as e.g. pressure in the vessel, high temperatures, corrosive or toxic products etc.

Take note of chapters "*Mounting*" and "*Connecting to voltage supply*" and carry out the listed steps in reverse order.

8.2 Disposal



Pass the instrument on to a specialised recycling company and do not use the municipal collecting points.

Remove any batteries in advance, if they can be removed from the device, and dispose of them separately.

If personal data is stored on the old device to be disposed of, delete it before disposal.

If you have no way to dispose of the old instrument properly, please contact us concerning return and disposal.

9 Supplement

9.1 Technical data

Note for approved instruments

The technical data in the respective safety instructions which are included in delivery are valid for approved instruments (e.g. with Ex approval). These data can differ from the data listed herein, for example regarding the process conditions or the voltage supply.

All approval documents can be downloaded from our homepage.

General data

Materials 316Ti corresponds to 1.4571, 316L corresponds to 1.4404 or 1.4435

Materials, wetted parts

- | | |
|----------------------------|---------------------------|
| – Process fitting - thread | PP (Polypropylene) |
| – Electrode | 316Ti, Alloy C22 (2.4602) |
| – Process seal | Klingsil C-4400 |

Materials, non-wetted parts

- | | |
|---|--|
| – Housing | Plastic PBT (polyester), Alu die-casting powder-coated |
| – Seal ring between housing and housing cover | Silicone |
| – Ground terminal | 316L |
| – Cable gland | PA, stainless steel, brass |
| – Sealing, cable gland | NBR |
| – Blind plug, cable gland | PA |

Weights

- | | |
|--------------------------|---------------------|
| – With plastic housing | 550 g (19.4 oz) |
| – with Aluminium housing | 850 g (30 oz) |
| – Electrode | 100 g/m (1.1 oz/ft) |

Probe length (L)

- | | |
|--------|--------------------|
| – min. | 120 mm (4.7 in) |
| – max. | 4000 mm (157.5 in) |

Process fittings

- | | |
|-----------------------|-------------|
| – Thread (DIN 3852-A) | G1½ (PN 25) |
|-----------------------|-------------|

Measurement voltage

approx. 3 V_{eff}

Measurement current

< 3 mA

Output variable

Output

Relay output (DPDT), 2 floating spdts

Switching voltage

max. 253 V AC/DC

With circuits > 150 V AC/DC, the relay contacts must be in the same circuit.

Switching current

max. 3 A AC (cos phi > 0,9), 1 A DC

Breaking capacity

- Min. 50 mW
 - Max. 750 VA AC, 40 W DC (bei $U < 40$ V DC)
- If inductive loads or stronger currents are switched through, the gold plating on the relay contact surface will be permanently damaged. The contact is then no longer suitable for switching low-level signal circuits.

Contact material (relay contacts)

AgNi or AgSnO₂ each with 3 µm gold plating

Modes (switchable)

- A Max. detection or overflow protection
- B Min. detection or dry run protection

Switching delay

0.5 ... 20 s

Ambient conditions

Ambient temperature on the housing -40 ... +70 °C (-40 ... +158 °F)

Ambient temperature with operating voltage > 60 V DC -40 ... +50 °C (-40 ... +122 °F)

Storage and transport temperature -40 ... +80 °C (-40 ... +176 °F)

Process conditions

Permissible process temperature -40 ... +100 °C (-40 ... +212 °F)

Process pressure -1 ... 6 bar/-100 ... 600 kPa (-14.5 ... 87 psig)

Conductance of the medium min. 5 µS/cm with 30 mm electrode covering

Electromechanical data

Options of the cable entry

- Cable entry M20 x 1.5
- Cable gland M20 x 1.5
- Blind plug M20 x 1.5

Wire cross-section (screw terminals)

- Massive wire, stranded wire 0.2 ... 2.5 mm² (AWG 24 ... 14)
- Stranded wire with end sleeve 0.2 ... 1.5 mm² (AWG 24 ... 16)

Voltage supply

Operating voltage 20 ... 253 V AC, 50/60 Hz, 20 ... 72 V DC (at $U > 60$ V DC, the ambient temperature can be max. 50 °C/122 °F)

Power consumption 1 ... 9 VA (AC), approximately 1.5 W (DC)

Electrical protective measures

Protection rating

- Plastic housing IP66 (NEMA Type 4X)
- Aluminium housing IP66/IP 67 (NEMA Type 4X)

Pollution degree ¹⁾	4
Overvoltage category	II
Protection class	I

9.2 Dimensions




Fig. 8: VEGAKON 66 with three probes

The length of the electrode is calculated from the seal surface of the thread.
Hence the metallic rods are 92.5 mm (3.64 in) shorter.

L1 Length ground probe from seal surface

L2 Length max. probe from seal surface

L3 Length min. probe from seal surface

¹⁾ When used with fulfilled housing protection.

9.3 Industrial property rights

VEGA product lines are global protected by industrial property rights. Further information see www.vega.com.

VEGA Produktfamilien sind weltweit geschützt durch gewerbliche Schutzrechte.

Nähere Informationen unter www.vega.com.

Les lignes de produits VEGA sont globalement protégées par des droits de propriété intellectuelle. Pour plus d'informations, on pourra se référer au site www.vega.com.

VEGA lineas de productos están protegidas por los derechos en el campo de la propiedad industrial. Para mayor información revise la pagina web www.vega.com.

Линии продукции фирмы ВЕГА защищаются по всему миру правами на интеллектуальную собственность. Дальнейшую информацию смотрите на сайте www.vega.com.

VEGA系列产品在全球享有知识产权保护。

进一步信息请参见网站 < www.vega.com。

9.4 Trademark

All the brands as well as trade and company names used are property of their lawful proprietor/originator.

Printing date:

VEGA

For more VEGA product please visit: www.ainstru.com
Mail: Sales@ainstru.com

All statements concerning scope of delivery, application, practical use and operating conditions of the sensors and processing systems correspond to the information available at the time of printing.
Subject to change without prior notice

© VEGA Grieshaber KG, Schiltach/Germany 2022



32649-EN-220324

VEGA Grieshaber KG
Am Hohenstein 113
77761 Schiltach
Germany

Phone +49 7836 50-0
E-mail: info.de@vega.com
www.vega.com